

ATTACHMENT 2: DATA VALIDATION REPORTS (SUMMARY SHEETS)

This attachment contains the summary sheets from the data validation performed by an independent subcontractor, Laboratory Data Consultants, Inc. (LDC), Carlsbad, CA. Complete data validation reports are available upon request.

NASA JPL
Data Validation Reports

Volatiles

LDc

Laboratory Data Consultants, Inc.
Data Validation Report

Project/Site Name: NASA JPL
Collection Date: April 26, 2005
LDC Report Date: July 29, 2005
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level III & IV
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2457

Sample Identification

EB-1-4/26/05
MW-21-1
MW-21-2**
MW-21-3
MW-21-4
MW-21-5
TB-1-4/26/05
MW-21-2MS
MW-21-2MSD

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 9 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% .

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were not required by the method.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

All target compound identifications were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XII. Compound Quantitation and CRQLs

All compound quantitation and CRQLs were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XIII. Tentatively Identified Compounds (TICs)

Tentatively identified compounds were not reported by the laboratory.

XIV. System Performance

The system performance was within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XV. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

XVII. Field Blanks

Sample TB-1-4/26/05 was identified as a trip blank. No volatile contaminants were found in this blank.

Sample EB-1-4/26/05 was identified as an equipment blank. No volatile contaminants were found in this blank with the following exceptions:

Equipment Blank ID	Compound	Concentration (ug/L)
EB-1-4/26/05	2-Butanone	13
	Ethylbenzene	1.1
	Toluene	0.8
	o-Xylene	1.4
	m,p-Xylenes	5.5

NASA JPL

Volatiles - Data Qualification Summary - SDG 05-2457

No Sample Data Qualified in this SDG

NASA JPL

Volatiles - Laboratory Blank Data Qualification Summary - SDG 05-2457

No Sample Data Qualified in this SDG

Laboratory Data Consultants, Inc.
Data Validation Report

Project/Site Name: NASA JPL
Collection Date: April 27, 2005
LDC Report Date: August 1, 2005
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2485

Sample Identification

EB-2-4/27/05
MW-19-1
MW-19-2
MW-19-3
MW-19-4
MW-19-5
TB-2-4/27/05
MW-19-2MS
MW-19-2MSD
MW-19-3MS
MW-19-3MSD

Introduction

This data review covers 11 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990.

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% with the following exceptions:

Date	Compound	%D	Associated Samples	Flag	A or P
5/5/05	2,2-Dichloropropane	34.62	MW-19-2 MW-19-2MS MW-19-2MSD 05G1961MB01	J (all detects) UJ (all non-detects)	A

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

Raw data were not reviewed for this SDG.

XII. Compound Quantitation and CRQLs

Raw data were not reviewed for this SDG.

XIII. Tentatively Identified Compounds (TICs)

Raw data were not reviewed for this SDG.

XIV. System Performance

Raw data were not reviewed for this SDG.

XV. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

XVII. Field Blanks

Sample TB-2-4/27/05 was identified as a trip blank. No volatile contaminants were found in this blank.

Sample EB-2-4/27/05 was identified as an equipment blank. No volatile contaminants were found in this blank with the following exceptions:

Equipment Blank ID	Compound	Concentration (ug/L)
EB-2-4/27/05	Ethylbenzene	0.7
	Toluene	0.4
	o-Xylene	0.8
	m,p-Xylenes	3.8

NASA JPL**Volatiles - Data Qualification Summary - SDG 05-2485**

SDG	Sample	Compound	Flag	A or P	Reason
05-2485	MW-19-2	2,2-Dichloropropane	J (all detects) UJ (all non-detects)	A	Continuing calibration (%D)

NASA JPL**Volatiles - Laboratory Blank Data Qualification Summary - SDG 05-2485**

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: April 29, 2005
LDC Report Date: August 1, 2005
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2522

Sample Identification

EB-3-4/29/05
MW-20-1
MW-20-2
MW-20-3
MW-20-4
MW-20-5
TB-3-4/29/05
MW-20-4MS
MW-20-4MSD

Introduction

This data review covers 9 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% with the following exceptions:

Date	Compound	%D	Associated Samples	Flag	A or P
5/5/05	2,2-Dichloropropane	34.62	All samples in SDG 05-2522	J (all detects) UJ (all non-detects)	A

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were not required by the method.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

Raw data were not reviewed for this SDG.

XII. Compound Quantitation and CRQLs

Raw data were not reviewed for this SDG.

XIII. Tentatively Identified Compounds (TICs)

Raw data were not reviewed for this SDG.

XIV. System Performance

Raw data were not reviewed for this SDG.

XV. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

XVII. Field Blanks

Sample TB-3-4/29/05 was identified as a trip blank. No volatile contaminants were found in this blank.

Sample EB-3-4/29/05 was identified as an equipment blank. No volatile contaminants were found in this blank with the following exceptions:

Equipment Blank ID	Compound	Concentration (ug/L)
EB-3-4/29/05	o-Xylene m,p-Xylenes	0.3 1.5

NASA JPL**Volatiles - Data Qualification Summary - SDG 05-2522**

SDG	Sample	Compound	Flag	A or P	Reason
05-2522	EB-3-4/29/05 MW-20-1 MW-20-2 MW-20-3 MW-20-4 MW-20-5 TB-3-4/29/05	2,2-Dichloropropane	J (all detects) UJ (all non-detects)	A	Continuing calibration (%D)

NASA JPL**Volatiles - Laboratory Blank Data Qualification Summary - SDG 05-2522**

No Sample Data Qualified in this SDG

Laboratory Data Consultants, Inc.
Data Validation Report

Project/Site Name: NASA JPL
Collection Date: May 2, 2005
LDC Report Date: July 11, 2005
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level III & IV
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2545

Sample Identification

EB-4-5/2/05
MW-17-1
MW-17-2
MW-17-3
MW-17-4**
MW-17-5
TB-4-5/2/05

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 7 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% with the following exceptions:

Date	Compound	%D	Associated Samples	Flag	A or P
5/5/05	2,2-Dichloropropane	34.62	All samples in SDG 05-2545	J (all detects) UJ (all non-detects)	P

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were not required by the method.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

All target compound identifications were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XII. Compound Quantitation and CRQLs

All compound quantitation and CRQLs were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XIII. Tentatively Identified Compounds (TICs)

Tentatively identified compounds were not reported by the laboratory.

XIV. System Performance

The system performance was within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XV. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

XVII. Field Blanks

Sample TB-4-5/2/05 was identified as a trip blank. No volatile contaminants were found in this blank.

Sample EB-4-5/2/05 was identified as an equipment blank. No volatile contaminants were found in this blank with the following exceptions:

Equipment Blank ID	Compound	Concentration (ug/L)
EB-4-5/2/05	Ethylbenzene	0.7
	Toluene	0.4
	o-Xylene	0.8
	m,p-Xylenes	4.1

NASA JPL**Volatiles - Data Qualification Summary - SDG 05-2545**

SDG	Sample	Compound	Flag	A or P	Reason
05-2545	EB-4-5/2/05 MW-17-1 MW-17-2 MW-17-3 MW-17-4** MW-17-5 TB-4-5/2/05	2,2-Dichloropropane	J (all detects) UJ (all non-detects)	P	Continuing calibration (%D)

NASA JPL**Volatiles - Laboratory Blank Data Qualification Summary - SDG 05-2545**

No Sample Data Qualified in this SDG

Laboratory Data Consultants, Inc.
Data Validation Report

Project/Site Name: NASA JPL
Collection Date: May 3, 2005
LDC Report Date: August 1, 2005
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2557

Sample Identification

DUPE-1-2Q05
EB-5-5/3/05
MW-18-1
MW-18-2
MW-18-3
MW-18-4
MW-18-5
TB-5-5/3/05

Introduction

This data review covers 8 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% .

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

Raw data were not reviewed for this SDG.

XII. Compound Quantitation and CRQLs

Raw data were not reviewed for this SDG.

XIII. Tentatively Identified Compounds (TICs)

Raw data were not reviewed for this SDG.

XIV. System Performance

Raw data were not reviewed for this SDG.

XV. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

Samples DUPE-1-2Q05 and MW-18-2 were identified as field duplicates. No volatiles were detected in any of the samples.

XVII. Field Blanks

Sample TB-5-5/3/05 was identified as a trip blank. No volatile contaminants were found in this blank.

Sample EB-5-5/3/05 was identified as an equipment blank. No volatile contaminants were found in this blank with the following exceptions:

Equipment Blank ID	Compound	Concentration (ug/L)
EB-5-5/3/05	Chloroform	0.3
	Ethylbenzene	0.3
	o-Xylene	0.3
	m,p-Xylenes	1.9

NASA JPL

Volatiles - Data Qualification Summary - SDG 05-2557

No Sample Data Qualified in this SDG

NASA JPL

Volatiles - Laboratory Blank Data Qualification Summary - SDG 05-2557

No Sample Data Qualified in this SDG

Laboratory Data Consultants, Inc.
Data Validation Report

Project/Site Name: NASA JPL
Collection Date: May 4, 2005
LDC Report Date: August 1, 2005
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2586

Sample Identification

EB-6-5/4/05
MW-3-1
MW-3-2
MW-3-3
MW-3-4
MW-3-5
TB-6-5/4/05
MW-3-3MS
MW-3-3MSD

Introduction

This data review covers 9 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% .

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

Raw data were not reviewed for this SDG.

XII. Compound Quantitation and CRQLs

Raw data were not reviewed for this SDG.

XIII. Tentatively Identified Compounds (TICs)

Raw data were not reviewed for this SDG.

XIV. System Performance

Raw data were not reviewed for this SDG.

XV. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

XVII. Field Blanks

Sample TB-6-5/4/05 was identified as a trip blank. No volatile contaminants were found in this blank.

Sample EB-6-5/4/05 was identified as an equipment blank. No volatile contaminants were found in this blank with the following exceptions:

Equipment Blank ID	Compound	Concentration (ug/L)
EB-6-5/4/05	m,p-Xylenes	1.7

NASA JPL

Volatiles - Data Qualification Summary - SDG 05-2586

No Sample Data Qualified in this SDG

NASA JPL

Volatiles - Laboratory Blank Data Qualification Summary - SDG 05-2586

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 6, 2005
LDC Report Date: July 29, 2005
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level III & IV
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2625

Sample Identification

DUPE-2-2Q05
DUPE-3-2Q05
MW-1**
MW-9
MW-15**
TB-7-5/6/05
MW-15MS
MW-15MSD

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 8 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% with the following exceptions:

Date	Compound	%D	Associated Samples	Flag	A or P
5/9/05	Chloroethane	33.39	All samples in SDG 05-2625	J (all detects) UJ (all non-detects)	A

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

All target compound identifications were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XII. Compound Quantitation and CRQLs

All compound quantitation and CRQLs were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XIII. Tentatively Identified Compounds (TICs)

Tentatively identified compounds were not reported by the laboratory.

XIV. System Performance

The system performance was within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XV. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

Samples DUPE-2-2Q05 and MW-1** and samples DUPE-3-2Q05 and MW-9 were identified as field duplicates. No volatiles were detected in any of the samples.

XVII. Field Blanks

Sample TB-7-5/6/05 was identified as a trip blank. No volatile contaminants were found in this blank.

NASA JPL**Volatiles - Data Qualification Summary - SDG 05-2625**

SDG	Sample	Compound	Flag	A or P	Reason
05-2625	DUPE-2-2Q05 DUPE-3-2Q05 MW-1** MW-9 MW-15** TB-7-5/6/05	Chloroethane	J (all detects) UJ (all non-detects)	A	Continuing calibration (%D)

NASA JPL**Volatiles - Laboratory Blank Data Qualification Summary - SDG 05-2625**

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 9, 2005
LDC Report Date: July 11, 2005
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 05-2652

Sample Identification

DUPE-4-2Q05
EB-7-5/9/05
MW-14-1
MW-14-2
MW-14-3
MW-14-4
MW-14-5
TB-8-5/9/05
DUPE-4-2Q05MS
DUPE-4-2Q05MSD

Introduction

This data review covers 10 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% .

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

Raw data were not reviewed for this SDG.

XII. Compound Quantitation and CRQLs

Raw data were not reviewed for this SDG.

XIII. Tentatively Identified Compounds (TICs)

Raw data were not reviewed for this SDG.

XIV. System Performance

Raw data were not reviewed for this SDG.

XV. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

Samples DUPE-4-2Q05 and MW-14-4 were identified as field duplicates. No volatiles were detected in any of the samples with the following exceptions:

Compound	Concentration (ug/L)		RPD
	DUPE-4-2Q05	MW-14-4	
2-Butanone	0.9	10U	200

XVII. Field Blanks

Sample TB-8-5/9/05 was identified as a trip blank. No volatile contaminants were found in this blank.

Sample EB-7-5/9/05 was identified as an equipment blank. No volatile contaminants were found in this blank with the following exceptions:

Equipment Blank ID	Compound	Concentration (ug/L)
EB-7-5/9/05	2-Butanone	0.7

NASA JPL

Volatiles - Data Qualification Summary - SDG 05-2652

No Sample Data Qualified in this SDG

NASA JPL

Volatiles - Laboratory Blank Data Qualification Summary - SDG 05-2652

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 10, 2005
LDC Report Date: July 11, 2005
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 05-2660

Sample Identification

DUPE-5-2Q05
EB-8-5/10/05
MW-22-1
MW-22-2
MW-22-3
MW-22-4
MW-22-5
TB-9-5/10/05

Introduction

This data review covers 8 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% .

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were not required by the method.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

Raw data were not reviewed for this SDG.

XII. Compound Quantitation and CRQLs

Raw data were not reviewed for this SDG.

XIII. Tentatively Identified Compounds (TICs)

Raw data were not reviewed for this SDG.

XIV. System Performance

Raw data were not reviewed for this SDG.

XV. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

Samples DUPE-5-2Q05 and MW-22-3 were identified as field duplicates. No volatiles were detected in any of the samples.

XVII. Field Blanks

Sample TB-9-5/10/05 was identified as a trip blank. No volatile contaminants were found in this blank.

Sample EB-8-5/10/05 was identified as an equipment blank. No volatile contaminants were found in this blank with the following exceptions:

Equipment Blank ID	Compound	Concentration (ug/L)
EB-8-5/10/05	Ethylbenzene	0.4
	o-Xylene	0.4
	m,p-Xylenes	2.1

NASA JPL

Volatiles - Data Qualification Summary - SDG 05-2660

No Sample Data Qualified in this SDG

NASA JPL

Volatiles - Laboratory Blank Data Qualification Summary - SDG 05-2660

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 11, 2005
LDC Report Date: July 11, 2005
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 05-2679

Sample Identification

EB-9-5/11/05
MW-24-1
MW-24-2
MW-24-3
MW-24-4
MW-24-5
TB-9-5/11/05

Introduction

This data review covers 7 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% with the following exceptions:

Date	Compound	%D	Associated Samples	Flag	A or P
5/18/05	Bromomethane	36.98	All samples in SDG 05-2679	J (all detects)	P
	2,2-Dichloropropane	36.61		UJ (all non-detects) J (all detects) UJ (all non-detects)	

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were not required by the method.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

Raw data were not reviewed for this SDG.

XII. Compound Quantitation and CRQLs

Raw data were not reviewed for this SDG.

XIII. Tentatively Identified Compounds (TICs)

Raw data were not reviewed for this SDG.

XIV. System Performance

Raw data were not reviewed for this SDG.

XV. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

XVII. Field Blanks

Sample TB-9-5/11/05 was identified as a trip blank. No volatile contaminants were found in this blank.

Sample EB-9-5/11/05 was identified as an equipment blank. No volatile contaminants were found in this blank.

NASA JPL**Volatiles - Data Qualification Summary - SDG 05-2679**

SDG	Sample	Compound	Flag	A or P	Reason
05-2679	EB-9-5/11/05 MW-24-1 MW-24-2 MW-24-3 MW-24-4 MW-24-5 TB-9-5/11/05	Bromomethane 2,2-Dichloropropane	J (all detects) UJ (all non-detects) J (all detects) UJ (all non-detects)	P	Continuing calibration (%D)

NASA JPL**Volatiles - Laboratory Blank Data Qualification Summary - SDG 05-2679**

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 12, 2005
LDC Report Date: July 11, 2005
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2703

Sample Identification

DUPE-6-2Q05
EB-11-5/12/05
MW-25-1
MW-25-2
MW-25-3
MW-25-4
MW-25-5
TB-11-5/12/05
MW-25-1MS
MW-25-1MSD

Introduction

This data review covers 10 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% with the following exceptions:

Date	Compound	%D	Associated Samples	Flag	A or P
5/18/05	Bromomethane	36.98	All samples in SDG 05-2703	J (all detects)	P
	2,2-Dichloropropane	36.61		UJ (all non-detects) J (all detects) UJ (all non-detects)	

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

Raw data were not reviewed for this SDG.

XII. Compound Quantitation and CRQLs

Raw data were not reviewed for this SDG.

XIII. Tentatively Identified Compounds (TICs)

Raw data were not reviewed for this SDG.

XIV. System Performance

Raw data were not reviewed for this SDG.

XV. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

Samples DUPE-6-2Q05 and MW-25-2 were identified as field duplicates. No volatiles were detected in any of the samples.

XVII. Field Blanks

Sample TB-11-5/12/05 was identified as a trip blank. No volatile contaminants were found in this blank.

Sample EB-11-5/12/05 was identified as an equipment blank. No volatile contaminants were found in this blank with the following exceptions:

Equipment Blank ID	Compound	Concentration (ug/L)
EB-11-5/12/05	Ethylbenzene	0.4
	o-Xylene	0.5
	m,p-Xylenes	2.1

NASA JPL**Volatiles - Data Qualification Summary - SDG 05-2703**

SDG	Sample	Compound	Flag	A or P	Reason
05-2703	DUPE-6-2Q05 EB-11-5/12/05 MW-25-1 MW-25-2 MW-25-3 MW-25-4 MW-25-5 TB-11-5/12/05	Bromomethane 2,2-Dichloropropane	J (all detects) UJ (all non-detects) J (all detects) UJ (all non-detects)	P	Continuing calibration (%D)

NASA JPL**Volatiles - Laboratory Blank Data Qualification Summary - SDG 05-2703**

No Sample Data Qualified in this SDG

Laboratory Data Consultants, Inc.
Data Validation Report

Project/Site Name: NASA JPL
Collection Date: May 16, 2005
LDC Report Date: July 12, 2005
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 05-2738

Sample Identification

DUPE-7-2Q05
EB-11-5/16/05
MW-11-1
MW-11-2
MW-11-3
MW-11-4
MW-11-5
TB-12-5/16/05

Introduction

This data review covers 8 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% with the following exceptions:

Date	Compound	%D	Associated Samples	Flag	A or P
5/19/05	Bromomethane	30.21	All samples in SDG 05-2738	J (all detects)	P
	2,2-Dichloropropane	46.41		UJ (all non-detects) J (all detects) UJ (all non-detects)	

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were not required by the method.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

Raw data were not reviewed for this SDG.

XII. Compound Quantitation and CRQLs

Raw data were not reviewed for this SDG.

XIII. Tentatively Identified Compounds (TICs)

Raw data were not reviewed for this SDG.

XIV. System Performance

Raw data were not reviewed for this SDG.

XV. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

Samples DUPE-7-2Q05 and MW-11-3 were identified as field duplicates. No volatiles were detected in any of the samples.

XVII. Field Blanks

Sample TB-12-5/16/05 was identified as a trip blank. No volatile contaminants were found in this blank.

Sample EB-11-5/16/05 was identified as an equipment blank. No volatile contaminants were found in this blank with the following exceptions:

Equipment Blank ID	Compound	Concentration (ug/L)
EB-11-5/16/05	m,p-Xylenes	1.1

NASA JPL**Volatiles - Data Qualification Summary - SDG 05-2738**

SDG	Sample	Compound	Flag	A or P	Reason
05-2738	DUPE-7-2Q05 EB-11-5/16/05 MW-11-1 MW-11-2 MW-11-3 MW-11-4 MW-11-5 TB-12-5/16/05	Bromomethane 2,2-Dichloropropane	J (all detects) UJ (all non-detects) J (all detects) UJ (all non-detects)	P	Continuing calibration (%D)

NASA JPL**Volatiles - Laboratory Blank Data Qualification Summary - SDG 05-2738**

No Sample Data Qualified in this SDG

Laboratory Data Consultants, Inc.
Data Validation Report

Project/Site Name: NASA JPL
Collection Date: May 17, 2005
LDC Report Date: July 12, 2005
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 05-2754

Sample Identification

EB-12-5/17/05
MW-23-1
MW-23-2
MW-23-3
MW-23-4
MW-23-5
TB-13-5/17/05
MW-23-1MS
MW-23-1MSD

Introduction

This data review covers 9 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% with the following exceptions:

Date	Compound	%D	Associated Samples	Flag	A or P
5/19/05	Bromomethane	30.21	All samples in SDG 05-2754	J (all detects)	P
	2,2-Dichloropropane	46.41		UJ (all non-detects) J (all detects) UJ (all non-detects)	

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

Raw data were not reviewed for this SDG.

XII. Compound Quantitation and CRQLs

Raw data were not reviewed for this SDG.

XIII. Tentatively Identified Compounds (TICs)

Raw data were not reviewed for this SDG.

XIV. System Performance

Raw data were not reviewed for this SDG.

XV. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

XVII. Field Blanks

Sample TB-13-5/17/05 was identified as a trip blank. No volatile contaminants were found in this blank.

Sample EB-12-5/17/05 was identified as an equipment blank. No volatile contaminants were found in this blank.

NASA JPL**Volatiles - Data Qualification Summary - SDG 05-2754**

SDG	Sample	Compound	Flag	A or P	Reason
05-2754	EB-12-5/17/05 MW-23-1 MW-23-2 MW-23-3 MW-23-4 MW-23-5 TB-13-5/17/05	Bromomethane 2,2-Dichloropropane	J (all detects) UJ (all non-detects) J (all detects) UJ (all non-detects)	P	Continuing calibration (%D)

NASA JPL**Volatiles - Laboratory Blank Data Qualification Summary - SDG 05-2754**

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 18, 2005
LDC Report Date: July 12, 2005
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level III & IV
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2772

Sample Identification

EB-13-5/18/05
MW-12-1
MW-12-2
MW-12-3**
MW-12-4
MW-12-5
TB-14-5/18/05
MW-12-5MS
MW-12-5MSD

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 9 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% with the following exceptions:

Date	Compound	%D	Associated Samples	Flag	A or P
5/24/05	Bromoform	40.47	All samples in SDG 05-2772	J (all detects) UJ (all non-detects)	P

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

All target compound identifications were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XII. Compound Quantitation and CRQLs

All compound quantitation and CRQLs were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XIII. Tentatively Identified Compounds (TICs)

Tentatively identified compounds were not reported by the laboratory.

XIV. System Performance

The system performance was within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XV. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

XVII. Field Blanks

Sample TB-14-5/18/05 was identified as a trip blank. No volatile contaminants were found in this blank.

Sample EB-13-5/18/05 was identified as an equipment blank. No volatile contaminants were found in this blank.

NASA JPL**Volatiles - Data Qualification Summary - SDG 05-2772**

SDG	Sample	Compound	Flag	A or P	Reason
05-2772	EB-13-5/18/05 MW-12-1 MW-12-2 MW-12-3** MW-12-4 MW-12-5 TB-14-5/18/05	Bromoform	J (all detects) UJ (all non-detects)	P	Continuing calibration (%D)

NASA JPL**Volatiles - Laboratory Blank Data Qualification Summary - SDG 05-2772**

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 19, 2005
LDC Report Date: July 12, 2005
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level III & IV
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2788

Sample Identification

EB-14-5/19/05
MW-4-1**
MW-4-2
MW-4-3
MW-4-4
MW-4-5
TB-15-5/19/05

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 7 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% with the following exceptions:

Date	Compound	%D	Associated Samples	Flag	A or P
5/24/05	Bromoform	40.47	All samples in SDG 05-2788	J (all detects) UJ (all non-detects)	P

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were not required by the method.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

All target compound identifications were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XII. Compound Quantitation and CRQLs

All compound quantitation and CRQLs were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XIII. Tentatively Identified Compounds (TICs)

Tentatively identified compounds were not reported by the laboratory.

XIV. System Performance

The system performance was within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XV. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

XVII. Field Blanks

Sample TB-15-5/19/05 was identified as a trip blank. No volatile contaminants were found in this blank.

Sample EB-14-5/19/05 was identified as an equipment blank. No volatile contaminants were found in this blank with the following exceptions:

Equipment Blank ID	Compound	Concentration (ug/L)
EB-14-5/19/05	Chloroform	0.3

NASA JPL
Volatiles - Data Qualification Summary - SDG 05-2788

SDG	Sample	Compound	Flag	A or P	Reason
05-2788	EB-14-5/19/05 MW-4-1** MW-4-2 MW-4-3 MW-4-4 MW-4-5 TB-15-5/19/05	Bromoform	J (all detects) UJ (all non-detects)	P	Continuing calibration (%D)

NASA JPL
Volatiles - Laboratory Blank Data Qualification Summary - SDG 05-2788

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 20, 2005
LDC Report Date: July 12, 2005
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 05-2812

Sample Identification

DUPE-8-2Q05
MW-5
MW-6
MW-16
TB-16-5/20/05
MW-5MS
MW-5MSD

Introduction

This data review covers 7 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990.

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% with the following exceptions:

Date	Compound	%D	Associated Samples	Flag	A or P
5/25/05	Bromoform 1,2-Dibromo-3-chloropropane	51.45 34.14	All samples in SDG 05-2812	J (all detects) UJ (all non-detects) J (all detects) UJ (all non-detects)	P

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were not required by the method.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

Raw data were not reviewed for this SDG.

XII. Compound Quantitation and CRQLs

Raw data were not reviewed for this SDG.

XIII. Tentatively Identified Compounds (TICs)

Raw data were not reviewed for this SDG.

XIV. System Performance

Raw data were not reviewed for this SDG.

XV. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

Samples DUPE-8-2Q05 and MW-6 were identified as field duplicates. No volatiles were detected in any of the samples with the following exceptions:

Compound	Concentration (ug/L)		RPD
	DUPE-8-2Q05	MW-6	
Chloroform	0.4	0.4	0

Compound	Concentration (ug/L)		RPD
	DUPE-8-2Q05	MW-6	
1,1-Dichloroethane	0.7	0.7	0
Tetrachloroethene	2.2	2.1	5

XVII. Field Blanks

Sample TB-16-5/20/05 was identified as a trip blank. No volatile contaminants were found in this blank.

NASA JPL**Volatiles - Data Qualification Summary - SDG 05-2812**

SDG	Sample	Compound	Flag	A or P	Reason
05-2812	DUPE-8-2Q05 MW-5 MW-6 MW-16 TB-16-5/20/05	Bromoform 1,2-Dibromo-3-chloropropane	J (all detects) UJ (all non-detects) J (all detects) UJ (all non-detects)	P	Continuing calibration (%D)

NASA JPL**Volatiles - Laboratory Blank Data Qualification Summary - SDG 05-2812**

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 23, 2005
LDC Report Date: July 12, 2005
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level III & IV
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2820

Sample Identification

MW-13**
TB-17-5/23/05

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 2 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% with the following exceptions:

Date	Compound	%D	Associated Samples	Flag	A or P
5/25/05	Bromoform 1,2-Dibromo-3-chloropropane	51.45 34.14	All samples in SDG 05-2820	J (all detects) UJ (all non-detects) J (all detects) UJ (all non-detects)	P

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were not required by the method.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

All target compound identifications were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XII. Compound Quantitation and CRQLs

All compound quantitation and CRQLs were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XIII. Tentatively Identified Compounds (TICs)

Tentatively identified compounds were not reported by the laboratory.

XIV. System Performance

The system performance was within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XV. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

XVII. Field Blanks

Sample TB-17-5/23/05 was identified as a trip blank. No volatile contaminants were found in this blank.

NASA JPL**Volatiles - Data Qualification Summary - SDG 05-2820**

SDG	Sample	Compound	Flag	A or P	Reason
05-2820	MW-13** TB-17-5/23/05	Bromoform 1,2-Dibromo-3-chloropropane	J (all detects) UJ (all non-detects) J (all detects) UJ (all non-detects)	P	Continuing calibration (%D)

NASA JPL**Volatiles - Laboratory Blank Data Qualification Summary - SDG 05-2820**

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 24, 2005
LDC Report Date: July 12, 2005
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level III & IV
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2832

Sample Identification

DUPE-9-2Q05
MW-7
MW-8**
MW-10
TB-18-5/24/05
MW-7MS
MW-7MSD
MW-8MS
MW-8MSD

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 9 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% with the following exceptions:

Date	Compound	%D	Associated Samples	Flag	A or P
5/25/05	Bromoform	51.45	All samples in SDG 05-2832	J (all detects)	P
	1,2-Dibromo-3-chloropropane	34.14		UJ (all non-detects) J (all detects) UJ (all non-detects)	

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

All target compound identifications were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XII. Compound Quantitation and CRQLs

All compound quantitation and CRQLs were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XIII. Tentatively Identified Compounds (TICs)

Tentatively identified compounds were not reported by the laboratory.

XIV. System Performance

The system performance was within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XV. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

Samples DUPE-9-2Q05 and MW-10 were identified as field duplicates. No volatiles were detected in any of the samples with the following exceptions:

Compound	Concentration (ug/L)		RPD
	DUPE-9-2Q05	MW-10	
Bromodichloromethane	0.5	0.4	22
Chloroform	1.1	1.1	0
Tetrachloroethene	0.4	0.4	0
Trichloroethene	5.8	5.5	5

XVII. Field Blanks

Sample TB-18-5/24/05 was identified as a trip blank. No volatile contaminants were found in this blank.

NASA JPL
Volatiles - Data Qualification Summary - SDG 05-2832

SDG	Sample	Compound	Flag	A or P	Reason
05-2832	DUPE-9-2Q05 MW-7 MW-8** MW-10 TB-18-5/24/05	Bromoform 1,2-Dibromo-3-chloropropane	J (all detects) UJ (all non-detects) J (all detects) UJ (all non-detects)	P	Continuing calibration (%D)

NASA JPL
Volatiles - Laboratory Blank Data Qualification Summary - SDG 05-2832

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 25, 2005
LDC Report Date: July 13, 2005
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level III & IV
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 05-2848

Sample Identification

EB-19-5/25/05
MW-26-1
MW-26-2**
SB-1-2Q05
TB-19-5/25/05
MW-26-2MS
MW-26-2MSD

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 7 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% with the following exceptions:

Date	Compound	%D	Associated Samples	Flag	A or P
6/3/05	Bromoform	51.11	All samples in SDG 05-2848	J (all detects) UJ (all non-detects)	P

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

All target compound identifications were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XII. Compound Quantitation and CRQLs

All compound quantitation and CRQLs were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XIII. Tentatively Identified Compounds (TICs)

Tentatively identified compounds were not reported by the laboratory.

XIV. System Performance

The system performance was within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XV. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

XVII. Field Blanks

Sample TB-19-5/25/05 was identified as a trip blank. No volatile contaminants were found in this blank.

Sample EB-19-5/25/05 was identified as an equipment blank. No volatile contaminants were found in this blank.

Sample SB-1-2Q05 was identified as a source blank. No volatile contaminants were found in this blank with the following exceptions:

Source Blank ID	Compound	Concentration (ug/L)
SB-1-2Q05	Ethylbenzene	5.5
	Toluene	1.5
	o-Xylene	8.3
	m,p-Xylenes	25.0

NASA JPL**Volatiles - Data Qualification Summary - SDG 05-2848**

SDG	Sample	Compound	Flag	A or P	Reason
05-2848	EB-19-5/25/05 MW-26-1 MW-26-2** SB-1-2Q05 TB-19-5/25/05	Bromoform	J (all detects) UJ (all non-detects)	P	Continuing calibration (%D)

NASA JPL**Volatiles - Laboratory Blank Data Qualification Summary - SDG 05-2848**

No Sample Data Qualified in this SDG

NASA JPL
Data Validation Reports

Metals

INDC

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: April 26, 2005
LDC Report Date: August 1, 2005
Matrix: Water
Parameters: Metals
Validation Level: EPA Level III & IV
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 05-2457

Sample Identification

EB-1-4/26/05
MW-21-1
MW-21-2**
MW-21-3
MW-21-4
MW-21-5
MW-21-2MS
MW-21-2MSD
MW-21-2DUP

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 9 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Methods 200.7, 200.8, and 200.9 for Metals. The metals analyzed were Arsenic, Calcium, Chromium, Iron, Lead, Magnesium, Potassium, and Sodium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (October 2004) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blanks are summarized in Section III.

Field duplicates are summarized in Section XIII.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

An initial calibration was performed.

The frequency and analysis criteria of the initial calibration verification (ICV) and continuing calibration verification (CCV) were met.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the initial, continuing and preparation blanks with the following exceptions:

Method Blank ID	Analyte	Maximum Concentration	Associated Samples
PB (prep blank)	Potassium	60.1 ug/L	All samples in SDG 05-2457
ICB/CCB	Arsenic Calcium Potassium	3.90 ug/L 651.869 ug/L 63.506 ug/L	All samples in SDG 05-2457

Data qualification by the initial, continuing and preparation blanks (ICB/CCB/PBs) was based on the maximum contaminant concentration in the ICB/CCB/PBs in the analysis of each analyte. The sample concentrations were either not detected or were significantly greater (>5X blank contaminants) than the concentrations found in the associated method blanks with the following exceptions:

Sample	Analyte	Reported Concentration	Modified Final Concentration
EB-1-4/26/05	Arsenic Calcium Potassium	3.0 ug/L 287 ug/L 67.5 ug/L	3.0U ug/L 287U ug/L 67.5U ug/L
MW-21-1	Arsenic	2.7 ug/L	2.7U ug/L
MW-21-3	Arsenic	4.2 ug/L	4.2U ug/L
MW-21-4	Arsenic	3.5 ug/L	3.5U ug/L

Sample	Analyte	Reported Concentration	Modified Final Concentration
MW-21-5	Arsenic	2.1 ug/L	2.1U ug/L

IV. ICP Interference Check Sample (ICS) Analysis

ICP interference check sample analysis was not required.

V. Matrix Spike Analysis

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VI. Duplicate Sample Analysis

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Internal Standards

All internal standard percent recoveries (%R) were within QC limits for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

IX. Furnace Atomic Absorption QC

All graphite furnace atomic absorption QC were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for samples reviewed by Level III criteria.

X. ICP Serial Dilution

ICP serial dilution analysis was performed by the laboratory. The analysis criteria were met with the following exceptions:

Diluted Sample	Analyte	%D (Limits)	Associated Samples	Flag	A or P
MW-21-2**L	Chromium	11.7 (≤ 10)	All samples in SDG 05-2457	J (all detects)	A

XI. Sample Result Verification

All sample result verifications were acceptable for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XIII. Field Duplicates

No field duplicates were identified in this SDG.

XIV. Field Blanks

Sample EB-1-4/26/05 was identified as an equipment blank. No metal contaminants were found in this blank with the following exceptions:

Equipment Blank ID	Analyte	Concentration (ug/L)
EB-1-4/26/05	Chromium	1.8
	Lead	0.027
	Arsenic	3.0
	Calcium	287
	Potassium	67.5

NASA JPL**Metals - Data Qualification Summary - SDG 05-2457**

SDG	Sample	Analyte	Flag	A or P	Reason
05-2457	EB-1-4/26/05 MW-21-1 MW-21-2** MW-21-3 MW-21-4 MW-21-5	Chromium	J (all detects)	A	ICP serial dilution (%D)

NASA JPL**Metals - Laboratory Blank Data Qualification Summary - SDG 05-2457**

SDG	Sample	Analyte	Modified Final Concentration	A or P
05-2457	EB-1-4/26/05	Arsenic Calcium Potassium	3.0U ug/L 287U ug/L 67.5U ug/L	A
05-2457	MW-21-1	Arsenic	2.7U ug/L	A
05-2457	MW-21-3	Arsenic	4.2U ug/L	A
05-2457	MW-21-4	Arsenic	3.5U ug/L	A
05-2457	MW-21-5	Arsenic	2.1U ug/L	A

Laboratory Data Consultants, Inc.
Data Validation Report

Project/Site Name: NASA JPL
Collection Date: April 27, 2005
LDC Report Date: August 1, 2005
Matrix: Water
Parameters: Metals
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2485

Sample Identification

EB-2-4/27/05
MW-19-1
MW-19-2
MW-19-3
MW-19-4
MW-19-5
MW-19-2MS
MW-19-2MSD

Introduction

This data review covers 8 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Methods 200.7, 200.8, and 200.9 for Metals. The metals analyzed were Arsenic, Calcium, Chromium, Iron, Lead, Magnesium, Potassium, and Sodium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (October 2004) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blanks are summarized in Section III.

Field duplicates are summarized in Section XIII.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

An initial calibration was performed.

The frequency and analysis criteria of the initial calibration verification (ICV) and continuing calibration verification (CCV) were met.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the initial, continuing and preparation blanks with the following exceptions:

Method Blank ID	Analyte	Maximum Concentration	Associated Samples
PB (prep blank)	Potassium	60.1 ug/L	All samples in SDG 05-2485
ICB/CCB	Arsenic Calcium Potassium	4.30 ug/L 651.869 ug/L 63.506 ug/L	All samples in SDG 05-2485

Data qualification by the initial, continuing and preparation blanks (ICB/CCB/PBs) was based on the maximum contaminant concentration in the ICB/CCB/PBs in the analysis of each analyte. The sample concentrations were either not detected or were significantly greater (>5X blank contaminants) than the concentrations found in the associated method blanks with the following exceptions:

Sample	Analyte	Reported Concentration	Modified Final Concentration
EB-2-4/27/05	Arsenic Calcium Potassium	3.1 ug/L 240 ug/L 66.3 ug/L	3.1U ug/L 240U ug/L 66.3U ug/L
MW-19-1	Arsenic	1.7 ug/L	1.7U ug/L
MW-19-2	Arsenic	1.8 ug/L	1.8U ug/L
MW-19-3	Arsenic	4.3 ug/L	4.3U ug/L

Sample	Analyte	Reported Concentration	Modified Final Concentration
MW-19-4	Arsenic	3.1 ug/L	3.1U ug/L
MW-19-5	Arsenic	4.1 ug/L	4.1U ug/L

IV. ICP Interference Check Sample (ICS) Analysis

ICP interference check sample analysis was not required.

V. Matrix Spike Analysis

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VI. Duplicate Sample Analysis

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Internal Standards

ICP-MS was not reviewed for this SDG.

IX. Furnace Atomic Absorption QC

All graphite furnace atomic absorption QC were within validation criteria.

X. ICP Serial Dilution

ICP serial dilution analysis was performed by the laboratory. The analysis criteria were met.

XI. Sample Result Verification

Raw data were not reviewed for this SDG.

XII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XIII. Field Duplicates

No field duplicates were identified in this SDG.

XIV. Field Blanks

Sample EB-2-4/27/05 was identified as an equipment blank. No metal contaminants were found in this blank with the following exceptions:

Equipment Blank ID	Analyte	Concentration (ug/L)
EB-2-4/27/05	Chromium	3.7
	Lead	0.014
	Arsenic	3.1
	Calcium	240
	Iron	9.3
	Magnesium	7.2
	Potassium	66.3

NASA JPL**Metals - Data Qualification Summary - SDG 05-2485**

No Sample Data Qualified in this SDG

NASA JPL**Metals - Laboratory Blank Data Qualification Summary - SDG 05-2485**

SDG	Sample	Analyte	Modified Final Concentration	A or P
05-2485	EB-2-4/27/05	Arsenic Calcium Potassium	3.1U ug/L 240U ug/L 66.3U ug/L	A
05-2485	MW-19-1	Arsenic	1.7U ug/L	A
05-2485	MW-19-2	Arsenic	1.8U ug/L	A
05-2485	MW-19-3	Arsenic	4.3U ug/L	A
05-2485	MW-19-4	Arsenic	3.1U ug/L	A
05-2485	MW-19-5	Arsenic	4.1U ug/L	A

Laboratory Data Consultants, Inc.
Data Validation Report

Project/Site Name: NASA JPL
Collection Date: April 29, 2005
LDC Report Date: August 1, 2005
Matrix: Water
Parameters: Metals
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2522

Sample Identification

EB-3-4/29/05
MW-20-1
MW-20-2
MW-20-3
MW-20-4
MW-20-5
MW-20-4MS
MW-20-4MSD
MW-20-4DUP

Introduction

This data review covers 9 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Methods 200.7, 200.8, and 200.9 for Metals. The metals analyzed were Arsenic, Calcium, Chromium, Iron, Lead, Magnesium, Potassium, and Sodium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (October 2004) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blanks are summarized in Section III.

Field duplicates are summarized in Section XIII.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

An initial calibration was performed.

The frequency and analysis criteria of the initial calibration verification (ICV) and continuing calibration verification (CCV) were met.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the initial, continuing and preparation blanks with the following exceptions:

Method Blank ID	Analyte	Maximum Concentration	Associated Samples
PB (prep blank)	Potassium	112 ug/L	All samples in SDG 05-2522
ICB/CCB	Calcium Iron Magnesium Potassium Sodium	182.849 ug/L 7.940 ug/L 20.053 ug/L 194.992 ug/L 452.043 ug/L	All samples in SDG 05-2522

Data qualification by the initial, continuing and preparation blanks (ICB/CCB/PBs) was based on the maximum contaminant concentration in the ICB/CCB/PBs in the analysis of each analyte. The sample concentrations were either not detected or were significantly greater (>5X blank contaminants) than the concentrations found in the associated method blanks with the following exceptions:

Sample	Analyte	Reported Concentration	Modified Final Concentration
EB-3-4/29/05	Iron Magnesium Potassium Sodium	10.5 ug/L 8.7 ug/L 108 ug/L 695 ug/L	10.5U ug/L 8.7U ug/L 108U ug/L 695U ug/L
MW-20-2	Iron	14.6 ug/L	14.6U ug/L
MW-20-3	Iron	16.0 ug/L	16.0U ug/L

IV. ICP Interference Check Sample (ICS) Analysis

ICP interference check sample analysis was not required.

V. Matrix Spike Analysis

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VI. Duplicate Sample Analysis

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Internal Standards

ICP-MS was not reviewed for this SDG.

IX. Furnace Atomic Absorption QC

All graphite furnace atomic absorption QC were within validation criteria.

X. ICP Serial Dilution

ICP serial dilution analysis was performed by the laboratory. The analysis criteria were met with the following exceptions:

Diluted Sample	Analyte	%D (Limits)	Associated Samples	Flag	A or P
MW-20-4L	Calcium	22.8 (≤ 10)	All samples in SDG 05-2522	J (all detects)	A

XI. Sample Result Verification

Raw data were not reviewed for this SDG.

XII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XIII. Field Duplicates

No field duplicates were identified in this SDG.

XIV. Field Blanks

Sample EB-3-4/29/05 was identified as an equipment blank. No metal contaminants were found in this blank with the following exceptions:

Equipment Blank ID	Analyte	Concentration (ug/L)
EB-3-4/29/05	Chromium	0.22
	Lead	0.023
	Iron	10.5
	Magnesium	8.7
	Potassium	108
	Sodium	695

NASA JPL**Metals - Data Qualification Summary - SDG 05-2522**

SDG	Sample	Analyte	Flag	A or P	Reason
05-2522	EB-3-4/29/05 MW-20-1 MW-20-2 MW-20-3 MW-20-4 MW-20-5	Calcium	J (all detects)	A	ICP serial dilution (%D)

NASA JPL**Metals - Laboratory Blank Data Qualification Summary - SDG 05-2522**

SDG	Sample	Analyte	Modified Final Concentration	A or P
05-2522	EB-3-4/29/05	Iron Magnesium Potassium Sodium	10.5U ug/L 8.7U ug/L 108U ug/L 695U ug/L	A
05-2522	MW-20-2	Iron	14.6U ug/L	A
05-2522	MW-20-3	Iron	16.0U ug/L	A

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 2, 2005
LDC Report Date: July 7, 2005
Matrix: Water
Parameters: Metals
Validation Level: EPA Level III & IV
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2545

Sample Identification

EB-4-5/2/05
MW-17-1
MW-17-2
MW-17-3
MW-17-4**
MW-17-5

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 6 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Methods 200.7, 200.8, and 200.9 for Metals. The metals analyzed were Arsenic, Calcium, Chromium, Iron, Lead, Magnesium, Potassium, and Sodium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blanks are summarized in Section III.

Field duplicates are summarized in Section XIII.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

An initial calibration was performed.

The frequency and analysis criteria of the initial calibration verification (ICV) and continuing calibration verification (CCV) were met.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the initial, continuing and preparation blanks with the following exceptions:

Method Blank ID	Analyte	Maximum Concentration	Associated Samples
PB (prep blank)	Potassium	112 ug/L	All samples in SDG 05-2545
ICB/CCB	Arsenic Calcium Iron Lead Magnesium Potassium Sodium	4.300 ug/L 182.849 ug/L 7.940 ug/L 0.021 ug/L 20.053 ug/L 194.992 ug/L 452.043 ug/L	All samples in SDG 05-2545

Data qualification by the initial, continuing and preparation blanks (ICB/CCB/PBs) was based on the maximum contaminant concentration in the ICB/CCB/PBs in the analysis of each analyte. The sample concentrations were either not detected or were significantly greater (>5X blank contaminants) than the concentrations found in the associated method blanks with the following exceptions:

Sample	Analyte	Reported Concentration	Modified Final Concentration
EB-4-5/2/05	Lead Magnesium Potassium	0.021 ug/L 6.6 ug/L 89.1 ug/L	0.021U ug/L 6.6U ug/L 89.1U ug/L
MW-17-1	Iron Lead	17.6 ug/L 0.023 ug/L	17.6U ug/L 0.023U ug/L

Sample	Analyte	Reported Concentration	Modified Final Concentration
MW-17-2	Lead	0.032 ug/L	0.032U ug/L
MW-17-3	Lead	0.097 ug/L	0.097U ug/L
MW-17-4**	Lead	0.052 ug/L	0.052U ug/L

IV. ICP Interference Check Sample (ICS) Analysis

The frequency of analysis was met.

The criteria for analysis were met.

V. Matrix Spike Analysis

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VI. Duplicate Sample Analysis

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Internal Standards

All internal standard percent recoveries (%R) were within QC limits for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

IX. Furnace Atomic Absorption QC

All graphite furnace atomic absorption QC were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for samples reviewed by Level III criteria.

X. ICP Serial Dilution

ICP serial dilution analysis was performed by the laboratory. The analysis criteria were met with the following exceptions:

Diluted Sample	Analyte	%D (Limits)	Associated Samples	Flag	A or P
MW-20-4L	Calcium	22.8 (≤ 10)	All samples in SDG 05-2545	J (all detects)	A
MW-21-2L	Chromium	11.7 (≤ 10)	All samples in SDG 05-2545	J (all detects)	A

XI. Sample Result Verification

All sample result verifications were acceptable for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XIII. Field Duplicates

No field duplicates were identified in this SDG.

XIV. Field Blanks

Sample EB-4-5/2/05 was identified as an equipment blank. No metal contaminants were found in this blank with the following exceptions:

Equipment Blank ID	Analyte	Concentration (ug/L)
EB-4-5/2/05	Chromium	0.37
	Lead	0.021
	Magnesium	6.6
	Potassium	89.1

NASA JPL

Metals - Data Qualification Summary - SDG 05-2545

SDG	Sample	Analyte	Flag	A or P	Reason
05-2545	EB-4-5/2/05 MW-17-1 MW-17-2 MW-17-3 MW-17-4** MW-17-5	Calcium Chromium	J (all detects) J (all detects)	A	ICP serial dilution (%D)

NASA JPL

Metals - Laboratory Blank Data Qualification Summary - SDG 05-2545

SDG	Sample	Analyte	Modified Final Concentration	A or P
05-2545	EB-4-5/2/05	Lead Magnesium Potassium	0.021U ug/L 6.6U ug/L 89.1U ug/L	A
05-2545	MW-17-1	Iron Lead	17.6U ug/L 0.023U ug/L	A
05-2545	MW-17-2	Lead	0.032U ug/L	A
05-2545	MW-17-3	Lead	0.097U ug/L	A
05-2545	MW-17-4**	Lead	0.052U ug/L	A

Laboratory Data Consultants, Inc.
Data Validation Report

Project/Site Name: NASA JPL
Collection Date: May 3, 2005
LDC Report Date: August 1, 2005
Matrix: Water
Parameters: Metals
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2557

Sample Identification

DUPE-1-2Q05
EB-5-5/3/05
MW-18-1
MW-18-2
MW-18-3
MW-18-4
MW-18-5
DUPE-1-2Q05MS
DUPE-1-2Q05MSD
DUPE-1-2Q05DUP

Introduction

This data review covers 10 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Methods 200.7, 200.8, and 200.9 for Metals. The metals analyzed were Arsenic, Calcium, Chromium, Iron, Lead, Magnesium, Potassium, and Sodium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (October 2004) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blanks are summarized in Section III.

Field duplicates are summarized in Section XIII.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

An initial calibration was performed.

The frequency and analysis criteria of the initial calibration verification (ICV) and continuing calibration verification (CCV) were met with the following exceptions:

Date	Lab. Reference/ID	Analyte	%R (Limits)	Associated Samples	Flag	A or P
5/4/05	CCV	Calcium Iron Magnesium Potassium Sodium	119.0 (90-110) 120.3 (90-110) 126.8 (90-110) 120.9 (90-110) 122.4 (90-110)	EB-5-5/3/05 MW-18-1 MW-18-2 MW-18-3 MW-18-4 MW-18-5	J (all detects) J (all detects) J (all detects) J (all detects) J (all detects)	P

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the initial, continuing and preparation blanks with the following exceptions:

Method Blank ID	Analyte	Maximum Concentration	Associated Samples
PB (prep blank)	Potassium	101 ug/L	All samples in SDG 05-2557
ICB/CCB	Arsenic Calcium Lead Potassium	4.30 ug/L 247.259 ug/L 0.021 ug/L 154.589 ug/L	All samples in SDG 05-2557

Data qualification by the initial, continuing and preparation blanks (ICB/CCB/PBs) was based on the maximum contaminant concentration in the ICB/CCB/PBs in the analysis of each analyte. The sample concentrations were either not detected or were significantly greater (>5X blank contaminants) than the concentrations found in the associated method blanks with the following exceptions:

Sample	Analyte	Reported Concentration	Modified Final Concentration
DUPE-1-2Q05	Arsenic Lead	3.7 ug/L 0.064 ug/L	3.7U ug/L 0.064U ug/L
EB-5-5/3/05	Arsenic Calcium Lead Potassium	3.5 ug/L 196 ug/L 0.061 ug/L 118 ug/L	3.5U ug/L 196U ug/L 0.061U ug/L 118U ug/L
MW-18-1	Arsenic Lead	5.9 ug/L 0.098 ug/L	5.9U ug/L 0.098U ug/L
MW-18-2	Arsenic Lead	4.4 ug/L 0.086 ug/L	4.4U ug/L 0.086U ug/L
MW-18-3	Arsenic Lead	6.5 ug/L 0.082 ug/L	6.5U ug/L 0.082U ug/L
MW-18-4	Arsenic Lead	3.6 ug/L 0.036 ug/L	3.6U ug/L 0.036U ug/L
MW-18-5	Arsenic Lead	3.6 ug/L 0.035 ug/L	3.6U ug/L 0.035U ug/L

IV. ICP Interference Check Sample (ICS) Analysis

ICP interference check sample analysis was not required.

V. Matrix Spike Analysis

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VI. Duplicate Sample Analysis

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Internal Standards

ICP-MS was not reviewed for this SDG.

IX. Furnace Atomic Absorption QC

All graphite furnace atomic absorption QC were within validation criteria.

X. ICP Serial Dilution

ICP serial dilution analysis was performed by the laboratory. The analysis criteria were met with the following exceptions:

Diluted Sample	Analyte	%D (Limits)	Associated Samples	Flag	A or P
DUPE-1-2Q05L	Iron	22.1 (≤ 10)	All samples in SDG 05-2557	J (all detects)	A
MW-21-2**L	Chromium	11.7 (≤ 10)	All samples in SDG 05-2557	J (all detects)	A

XI. Sample Result Verification

Raw data were not reviewed for this SDG.

XII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XIII. Field Duplicates

Samples DUPE-1-2Q05 and MW-18-2 were identified as field duplicates. No metals were detected in any of the samples with the following exceptions:

Compound	Concentration (ug/L)		RPD
	DUPE-1-2Q05	MW-18-2	
Arsenic	3.7	4.4	17
Calcium	53000	59100	11
Chromium	7.6	6.6	14
Iron	449	823	59
Lead	0.064	0.086	29

Compound	Concentration (ug/L)		RPD
	DUPE-1-2Q05	MW-18-2	
Magnesium	18800	20800	10
Potassium	2480	2690	8
Sodium	18700	20700	10

XIV. Field Blanks

Sample EB-5-5/3/05 was identified as an equipment blank. No metal contaminants were found in this blank with the following exceptions:

Equipment Blank ID	Analyte	Concentration (ug/L)
EB-5-5/3/05	Chromium	2.1
	Lead	0.061
	Arsenic	3.5
	Calcium	196
	Magnesium	6.1
	Potassium	118

NASA JPL
Metals - Data Qualification Summary - SDG 05-2557

SDG	Sample	Analyte	Flag	A or P	Reason
05-2557	EB-5-5/3/05 MW-18-1 MW-18-2 MW-18-3 MW-18-4 MW-18-5	Calcium Iron Magnesium Potassium Sodium	J (all detects) J (all detects) J (all detects) J (all detects) J (all detects)	P	Calibration (%R)
05-2557	DUPE-1-2Q05 EB-5-5/3/05 MW-18-1 MW-18-2 MW-18-3 MW-18-4 MW-18-5	Iron Chromium	J (all detects) J (all detects)	A	ICP serial dilution (%D)

NASA JPL
Metals - Laboratory Blank Data Qualification Summary - SDG 05-2557

SDG	Sample	Analyte	Modified Final Concentration	A or P
05-2557	DUPE-1-2Q05	Arsenic Lead	3.7U ug/L 0.064U ug/L	A
05-2557	EB-5-5/3/05	Arsenic Calcium Lead Potassium	3.5U ug/L 196U ug/L 0.061U ug/L 118U ug/L	A
05-2557	MW-18-1	Arsenic Lead	5.9U ug/L 0.098U ug/L	A
05-2557	MW-18-2	Arsenic Lead	4.4U ug/L 0.086U ug/L	A
05-2557	MW-18-3	Arsenic Lead	6.5U ug/L 0.082U ug/L	A
05-2557	MW-18-4	Arsenic Lead	3.6U ug/L 0.036U ug/L	A
05-2557	MW-18-5	Arsenic Lead	3.6U ug/L 0.035U ug/L	A

Laboratory Data Consultants, Inc.
Data Validation Report

Project/Site Name: NASA JPL
Collection Date: May 4, 2005
LDC Report Date: August 1, 2005
Matrix: Water
Parameters: Metals
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2586

Sample Identification

EB-6-5/4/05
MW-3-1
MW-3-2
MW-3-3
MW-3-4
MW-3-5
MW-3-3MS
MW-3-3MSD
MW-3-3DUP

Introduction

This data review covers 9 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Methods 200.7, 200.8, and 200.9 for Metals. The metals analyzed were Arsenic, Calcium, Chromium, Iron, Lead, Magnesium, Potassium, and Sodium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (October 2004) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blanks are summarized in Section III.

Field duplicates are summarized in Section XIII.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

An initial calibration was performed.

The frequency and analysis criteria of the initial calibration verification (ICV) and continuing calibration verification (CCV) were met.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the initial, continuing and preparation blanks with the following exceptions:

Method Blank ID	Analyte	Maximum Concentration	Associated Samples
PB (prep blank)	Potassium	93.5 ug/L	All samples in SDG 05-2586
ICB/CCB	Arsenic Calcium Iron Magnesium Potassium Sodium	1.40 ug/L 369.151 ug/L 6.157 ug/L 93.733 ug/L 152.346 ug/L 431.544 ug/L	All samples in SDG 05-2586

Data qualification by the initial, continuing and preparation blanks (ICB/CCB/PBs) was based on the maximum contaminant concentration in the ICB/CCB/PBs in the analysis of each analyte. The sample concentrations were either not detected or were significantly greater (>5X blank contaminants) than the concentrations found in the associated method blanks with the following exceptions:

Sample	Analyte	Reported Concentration	Modified Final Concentration
EB-6-5/4/05	Calcium Iron Magnesium Potassium Sodium	320 ug/L 18.1 ug/L 17.8 ug/L 128 ug/L 732 ug/L	320U ug/L 18.1U ug/L 17.8U ug/L 128U ug/L 732U ug/L
MW-3-1	Arsenic	1.5 ug/L	1.5U ug/L

Sample	Analyte	Reported Concentration	Modified Final Concentration
MW-3-3	Arsenic	1.1 ug/L	1.1U ug/L
MW-3-4	Arsenic	2.0 ug/L	2.0U ug/L
MW-3-5	Arsenic	2.1 ug/L	2.1U ug/L

IV. ICP Interference Check Sample (ICS) Analysis

ICP interference check sample analysis was not required.

V. Matrix Spike Analysis

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VI. Duplicate Sample Analysis

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Internal Standards

ICP-MS was not reviewed for this SDG.

IX. Furnace Atomic Absorption QC

All graphite furnace atomic absorption QC were within validation criteria.

X. ICP Serial Dilution

ICP serial dilution analysis was performed by the laboratory. The analysis criteria were met.

XI. Sample Result Verification

Raw data were not reviewed for this SDG.

XII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XIII. Field Duplicates

No field duplicates were identified in this SDG.

XIV. Field Blanks

Sample EB-6-5/4/05 was identified as an equipment blank. No metal contaminants were found in this blank with the following exceptions:

Equipment Blank ID	Analyte	Concentration (ug/L)
EB-6-5/4/05	Chromium	1.1
	Lead	0.019
	Calcium	320
	Iron	18.1
	Magnesium	17.8
	Potassium	128
	Sodium	732

NASA JPL**Metals - Data Qualification Summary - SDG 05-2586**

No Sample Data Qualified in this SDG

NASA JPL**Metals - Laboratory Blank Data Qualification Summary - SDG 05-2586**

SDG	Sample	Analyte	Modified Final Concentration	A or P
05-2586	EB-6-5/4/05	Calcium Iron Magnesium Potassium Sodium	320U ug/L 18.1U ug/L 17.8U ug/L 128U ug/L 732U ug/L	A
05-2586	MW-3-1	Arsenic	1.5U ug/L	A
05-2586	MW-3-3	Arsenic	1.1U ug/L	A
05-2586	MW-3-4	Arsenic	2.0U ug/L	A
05-2586	MW-3-5	Arsenic	2.1U ug/L	A

Laboratory Data Consultants, Inc.
Data Validation Report

Project/Site Name: NASA JPL
Collection Date: May 6, 2005
LDC Report Date: August 1, 2005
Matrix: Water
Parameters: Metals
Validation Level: EPA Level III & IV
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2625

Sample Identification

DUPE-2-2Q05
DUPE-3-2Q05
MW-1**
MW-9
MW-15**
MW-15MS
MW-15MSD
MW-15DUP

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 8 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Methods 200.7, 200.8, and 200.9 for Metals. The metals analyzed were Arsenic, Calcium, Chromium, Iron, Lead, Magnesium, Potassium, and Sodium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (October 2004) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blanks are summarized in Section III.

Field duplicates are summarized in Section XIII.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

An initial calibration was performed.

The frequency and analysis criteria of the initial calibration verification (ICV) and continuing calibration verification (CCV) were met.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the initial, continuing and preparation blanks with the following exceptions:

Method Blank ID	Analyte	Maximum Concentration	Associated Samples
PB (prep blank)	Calcium Potassium	156 ug/L 68.7 ug/L	All samples in SDG 05-2625
ICB/CCB	Arsenic Calcium Magnesium Potassium	1.40 ug/L 461.156 ug/L 7.943 ug/L 98.857 ug/L	All samples in SDG 05-2625

Data qualification by the initial, continuing and preparation blanks (ICB/CCB/PBs) was based on the maximum contaminant concentration in the ICB/CCB/PBs in the analysis of each analyte. The sample concentrations were either not detected or were significantly greater (>5X blank contaminants) than the concentrations found in the associated method blanks with the following exceptions:

Sample	Analyte	Reported Concentration	Modified Final Concentration
MW-1**	Arsenic	1.6 ug/L	1.6U ug/L
MW-9	Arsenic	1.2 ug/L	1.2U ug/L
MW-15**	Arsenic	1.5 ug/L	1.5U ug/L

IV. ICP Interference Check Sample (ICS) Analysis

ICP interference check sample analysis was not required.

V. Matrix Spike Analysis

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VI. Duplicate Sample Analysis

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Internal Standards

All internal standard percent recoveries (%R) were within QC limits for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

IX. Furnace Atomic Absorption QC

All graphite furnace atomic absorption QC were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for samples reviewed by Level III criteria.

X. ICP Serial Dilution

ICP serial dilution analysis was performed by the laboratory. The analysis criteria were met with the following exceptions:

Diluted Sample	Analyte	%D (Limits)	Associated Samples	Flag	A or P
MW-15**L	Sodium	11.3 (≤ 10)	All samples in SDG 05-2625	J (all detects)	A

XI. Sample Result Verification

All sample result verifications were acceptable for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XIII. Field Duplicates

Samples DUPE-2-2Q05 and MW-1** and samples DUPE-3-2Q05 and MW-9 were identified as field duplicates. No metals were detected in any of the samples with the following exceptions:

Compound	Concentration (ug/L)		RPD
	DUPE-2-2Q05	MW-1**	
Arsenic	5U	1.6	200
Calcium	67400	66900	1
Chromium	6.7	6.0	11
Iron	4.5	14.5	105
Lead	0.26	0.26	0
Magnesium	21600	20900	3
Potassium	3740	3630	3
Sodium	31000	30600	1

Compound	Concentration (ug/L)		RPD
	DUPE-3-2Q05	MW-9	
Arsenic	5U	1.2	200
Calcium	52300	50400	4
Chromium	2.1	2.3	9
Iron	131	147	12
Lead	0.55	0.65	17
Magnesium	16200	16000	1

Compound	Concentration (ug/L)		RPD
	DUPE-3-2Q05	MW-9	
Potassium	3100	2940	5
Sodium	22000	22200	1

XIV. Field Blanks

No field blanks were identified in this SDG.

NASA JPL**Metals - Data Qualification Summary - SDG 05-2625**

SDG	Sample	Analyte	Flag	A or P	Reason
05-2625	DUPE-2-2Q05 DUPE-3-2Q05 MW-1** MW-9 MW-15**	Iron	J (all detects)	A	ICP serial dilution (%D)

NASA JPL**Metals - Laboratory Blank Data Qualification Summary - SDG 05-2625**

SDG	Sample	Analyte	Modified Final Concentration	A or P
05-2625	MW-1**	Arsenic	1.6U ug/L	A
05-2625	MW-9	Arsenic	1.2U ug/L	A
05-2625	MW-15**	Arsenic	1.5U ug/L	A

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 9, 2005
LDC Report Date: July 6, 2005
Matrix: _____ Water _____
Parameters: Metals
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2652

Sample Identification

DUPE-4-2Q05
EB-7-5/9/05
MW-14-1
MW-14-2
MW-14-3
MW-14-4
MW-14-5

Introduction

This data review covers 7 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Methods 200.7, 200.8, and 200.9 for Metals. The metals analyzed were Arsenic, Calcium, Chromium, Iron, Lead, Magnesium, Potassium, and Sodium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blanks are summarized in Section III.

Field duplicates are summarized in Section XIII.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

An initial calibration was performed.

The frequency and analysis criteria of the initial calibration verification (ICV) and continuing calibration verification (CCV) were met.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the initial, continuing and preparation blanks with the following exceptions:

Method Blank ID	Analyte	Maximum Concentration	Associated Samples
PB (prep blank)	Calcium Potassium	156 ug/L 68.7 ug/L	All samples in SDG 05-2652
ICB/CCB	Arsenic Calcium Magnesium Potassium	1.400 ug/L 461.156 ug/L 7.943 ug/L 98.857 ug/L	All samples in SDG 05-2652

Data qualification by the initial, continuing and preparation blanks (ICB/CCB/PBs) was based on the maximum contaminant concentration in the ICB/CCB/PBs in the analysis of each analyte. The sample concentrations were either not detected or were significantly greater (>5X blank contaminants) than the concentrations found in the associated method blanks with the following exceptions:

Sample	Analyte	Reported Concentration	Modified Final Concentration
EB-7-5/9/05	Calcium Magnesium Potassium	329 ug/L 14.9 ug/L 88.0 ug/L	329U ug/L 14.9U ug/L 88.0U ug/L
MW-14-1	Arsenic	1.8 ug/L	1.8U ug/L
MW-14-3	Arsenic	1.1 ug/L	1.1U ug/L

Sample	Analyte	Reported Concentration	Modified Final Concentration
MW-14-5	Arsenic	3.0 ug/L	3.0U ug/L

IV. ICP Interference Check Sample (ICS) Analysis

ICP interference check sample analysis was not required.

V. Matrix Spike Analysis

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits with the following exceptions:

Spike ID (Associated Samples)	Analyte	MS (%R) (Limits)	MSD (%R) (Limits)	RPD (Limits)	Flag	A or P
MW-15MS/MSD (All samples in SDG 05-2652)	Arsenic	126 (75-125)	128 (75-125)	-	J (all detects)	A

VI. Duplicate Sample Analysis

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Internal Standards

Raw data were not reviewed for this SDG.

IX. Furnace Atomic Absorption QC

Raw data were not reviewed for this SDG.

X. ICP Serial Dilution

ICP serial dilution analysis was performed by the laboratory. The analysis criteria were met with the following exceptions:

Diluted Sample	Analyte	%D (Limits)	Associated Samples	Flag	A or P
MW-15L	Sodium	11.3 (≤ 10)	All samples in SDG 05-2652	J (all detects)	A

XI. Sample Result Verification

Raw data were not reviewed for this SDG.

XII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XIII. Field Duplicates

Samples DUPE-4-2Q05 and MW-14-4 were identified as field duplicates. No metals were detected in any of the samples with the following exceptions:

Analyte	Concentration (ug/L)		RPD
	DUPE-4-2Q05	MW-14-4	
Calcium	58500	59600	2
Chromium	6.9	6.3	9
Iron	116	107	8
Lead	0.043	0.13	101
Magnesium	20500	21100	3
Potassium	2340	2340	0
Sodium	30800	31500	2

XIV. Field Blanks

Sample EB-7-5/9/05 was identified as an equipment blank. No metal contaminants were found in this blank with the following exceptions:

Equipment Blank ID	Analyte	Concentration (ug/L)
EB-7-5/9/05	Lead	0.020
	Chromium	2.1
	Calcium	329
	Iron	7.6
	Magnesium	14.9
	Potassium	88.0

NASA JPL

Metals - Data Qualification Summary - SDG 05-2652

SDG	Sample	Analyte	Flag	A or P	Reason
05-2652	DUPE-4-2Q05 EB-7-5/9/05 MW-14-1 MW-14-2 MW-14-3 MW-14-4 MW-14-5	Arsenic	J (all detects)	A	Matrix spike/Matrix spike duplicates (%R)
05-2652	DUPE-4-2Q05 EB-7-5/9/05 MW-14-1 MW-14-2 MW-14-3 MW-14-4 MW-14-5	Sodium	J (all detects)	A	ICP serial dilution (%D)

NASA JPL

Metals - Laboratory Blank Data Qualification Summary - SDG 05-2652

SDG	Sample	Analyte	Modified Final Concentration	A or P
05-2652	EB-7-5/9/05	Calcium Magnesium Potassium	329U ug/L 14.9U ug/L 88.0U ug/L	A
05-2652	MW-14-1	Arsenic	1.8U ug/L	A
05-2652	MW-14-3	Arsenic	1.1U ug/L	A
05-2652	MW-14-5	Arsenic	3.0U ug/L	A

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 10, 2005
LDC Report Date: July 6, 2005
Matrix: Water
Parameters: Metals
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 05-2660

Sample Identification

DUPE-5-2Q05
EB-8-5/10/05
MW-22-1
MW-22-2
MW-22-3
MW-22-4
MW-22-5
DUPE-5-2Q05MS
DUPE-5-2Q05MSD
DUPE-5-2Q05DUP

Introduction

This data review covers 10 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Methods 200.7, 200.8, and 200.9 for Metals. The metals analyzed were Arsenic, Calcium, Chromium, Iron, Lead, Magnesium, Potassium, and Sodium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blanks are summarized in Section III.

Field duplicates are summarized in Section XIII.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

An initial calibration was performed.

The frequency and analysis criteria of the initial calibration verification (ICV) and continuing calibration verification (CCV) were met.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the initial, continuing and preparation blanks with the following exceptions:

Method Blank ID	Analyte	Maximum Concentration	Associated Samples
PB (prep blank)	Calcium Magnesium Potassium	331 ug/L 21.3 ug/L 83.9 ug/L	All samples in SDG 05-2660
ICB/CCB	Calcium Magnesium Potassium	320.685 ug/L 54.680 ug/L 87.451 ug/L	All samples in SDG 05-2660

Data qualification by the initial, continuing and preparation blanks (ICB/CCB/PBs) was based on the maximum contaminant concentration in the ICB/CCB/PBs in the analysis of each analyte. The sample concentrations were either not detected or were significantly greater (>5X blank contaminants) than the concentrations found in the associated method blanks with the following exceptions:

Sample	Analyte	Reported Concentration	Modified Final Concentration
EB-8-5/10/05	Magnesium Potassium	16.7 ug/L 100 ug/L	16.7U ug/L 100U ug/L

IV. ICP Interference Check Sample (ICS) Analysis

The frequency of analysis was met.

The criteria for analysis were met.

V. Matrix Spike Analysis

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VI. Duplicate Sample Analysis

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Internal Standards

Raw data were not reviewed for this SDG.

IX. Furnace Atomic Absorption QC

Raw data were not reviewed for this SDG.

X. ICP Serial Dilution

ICP serial dilution analysis was performed by the laboratory. The analysis criteria were met with the following exceptions:

Diluted Sample	Analyte	%D (Limits)	Associated Samples	Flag	A or P
DUPE-5-2Q05L	Calcium Magnesium	13.5 (≤ 10) 10.2 (≤ 10)	All samples in SDG 05-2660	J (all detects) J (all detects)	A

XI. Sample Result Verification

Raw data were not reviewed for this SDG.

XII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XIII. Field Duplicates

Samples DUPE-5-2Q05 and MW-22-3 were identified as field duplicates. No metals were detected in any of the samples with the following exceptions:

Analyte	Concentration (ug/L)		RPD
	DUPE-5-2Q05	MW-22-3	
Calcium	49700	49100	1
Chromium	5.3	5.0	6
Lead	0.054	0.043	23
Magnesium	20200	20100	0
Potassium	2350	2360	0
Sodium	40900	41400	1

XIV. Field Blanks

Sample EB-8-5/10/05 was identified as an equipment blank. No metal contaminants were found in this blank with the following exceptions:

Equipment Blank ID	Analyte	Concentration (ug/L)
EB-8-5/10/05	Chromium	3.7
	Lead	0.056
	Magnesium	16.7
	Potassium	100

NASA JPL
Metals - Data Qualification Summary - SDG 05-2660

SDG	Sample	Analyte	Flag	A or P	Reason
05-2660	DUPE-5-2Q05 EB-8-5/10/05 MW-22-1 MW-22-2 MW-22-3 MW-22-4 MW-22-5	Calcium Magnesium	J (all detects) J (all detects)	A	ICP serial dilution (%D)

NASA JPL
Metals - Laboratory Blank Data Qualification Summary - SDG 05-2660

SDG	Sample	Analyte	Modified Final Concentration	A or P
05-2660	EB-8-5/10/05	Magnesium Potassium	16.7U ug/L 100U ug/L	A

Laboratory Data Consultants, Inc.
Data Validation Report

Project/Site Name: NASA JPL
Collection Date: May 11, 2005
LDC Report Date: July 6, 2005
Matrix: Water
Parameters: Metals
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 05-2679

Sample Identification

EB-9-5/11/05
MW-24-1
MW-24-2
MW-24-3
MW-24-4
MW-24-5

Introduction

This data review covers 6 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Methods 200.7, 200.8, and 200.9 for Metals. The metals analyzed were Arsenic, Calcium, Chromium, Iron, Lead, Magnesium, Potassium, and Sodium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blanks are summarized in Section III.

Field duplicates are summarized in Section XIII.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

An initial calibration was performed.

The frequency and analysis criteria of the initial calibration verification (ICV) and continuing calibration verification (CCV) were met.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the initial, continuing and preparation blanks with the following exceptions:

Method Blank ID	Analyte	Maximum Concentration	Associated Samples
PB (prep blank)	Calcium Magnesium Potassium	331 ug/L 21.3 ug/L 83.9 ug/L	All samples in SDG 05-2679
ICB/CCB	Calcium Magnesium Potassium	320.685 ug/L 54.680 ug/L 87.451 ug/L	All samples in SDG 05-2679

Data qualification by the initial, continuing and preparation blanks (ICB/CCB/PBs) was based on the maximum contaminant concentration in the ICB/CCB/PBs in the analysis of each analyte. The sample concentrations were either not detected or were significantly greater (>5X blank contaminants) than the concentrations found in the associated method blanks with the following exceptions:

Sample	Analyte	Reported Concentration	Modified Final Concentration
EB-9-5/11/05	Magnesium Potassium	18.5 ug/L 99.3 ug/L	18.5U ug/L 99.3U ug/L

IV. ICP Interference Check Sample (ICS) Analysis

The frequency of analysis was met.

The criteria for analysis were met.

V. Matrix Spike Analysis

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VI. Duplicate Sample Analysis

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Internal Standards

Raw data were not reviewed for this SDG.

IX. Furnace Atomic Absorption QC

Raw data were not reviewed for this SDG.

X. ICP Serial Dilution

ICP serial dilution analysis was performed by the laboratory. The analysis criteria were met with the following exceptions:

Diluted Sample	Analyte	%D (Limits)	Associated Samples	Flag	A or P
DUPE-5-2Q05L	Calcium Magnesium	13.5 (≤ 10) 10.2 (≤ 10)	All samples in SDG 05-2679	J (all detects) J (all detects)	A

XI. Sample Result Verification

Raw data were not reviewed for this SDG.

XII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XIII. Field Duplicates

No field duplicates were identified in this SDG.

XIV. Field Blanks

Sample EB-9-5/11/05 was identified as an equipment blank. No metal contaminants were found in this blank with the following exceptions:

Equipment Blank ID	Analyte	Concentration (ug/L)
EB-9-5/11/05	Chromium	2.5
	Lead	0.029
	Magnesium	18.5
	Potassium	99.3
	Sodium	464

NASA JPL**Metals - Data Qualification Summary - SDG 05-2679**

SDG	Sample	Analyte	Flag	A or P	Reason
05-2679	EB-9-5/11/05 MW-24-1 MW-24-2 MW-24-3 MW-24-4 MW-24-5	Calcium Magnesium	J (all detects) J (all detects)	A	ICP serial dilution (%D)

NASA JPL**Metals - Laboratory Blank Data Qualification Summary - SDG 05-2679**

SDG	Sample	Analyte	Modified Final Concentration	A or P
05-2679	EB-9-5/11/05	Magnesium Potassium	18.5U ug/L 99.3U ug/L	A

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 12, 2005
LDC Report Date: July 6, 2005
Matrix: Water
Parameters: Metals
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 05-2703

Sample Identification

DUPE-6-2Q05
EB-11-5/12/05
MW-25-1
MW-25-2
MW-25-3
MW-25-4
MW-25-5

Introduction

This data review covers 7 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Methods 200.7, 200.8, and 200.9 for Metals. The metals analyzed were Arsenic, Calcium, Chromium, Iron, Lead, Magnesium, Potassium, and Sodium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blanks are summarized in Section III.

Field duplicates are summarized in Section XIII.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

An initial calibration was performed.

The frequency and analysis criteria of the initial calibration verification (ICV) and continuing calibration verification (CCV) were met.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the initial, continuing and preparation blanks with the following exceptions:

Method Blank ID	Analyte	Maximum Concentration	Associated Samples
PB (prep blank)	Calcium Magnesium Potassium	331 ug/L 21.3 ug/L 83.9 ug/L	All samples in SDG 05-2703
ICB/CCB	Calcium Magnesium Potassium	320.685 ug/L 54.680 ug/L 87.451 ug/L	All samples in SDG 05-2703

Data qualification by the initial, continuing and preparation blanks (ICB/CCB/PBs) was based on the maximum contaminant concentration in the ICB/CCB/PBs in the analysis of each analyte. The sample concentrations were either not detected or were significantly greater (>5X blank contaminants) than the concentrations found in the associated method blanks with the following exceptions:

Sample	Analyte	Reported Concentration	Modified Final Concentration
EB-11-5/12/05	Magnesium Potassium	26.2 ug/L 110 ug/L	26.2U ug/L 110U ug/L

IV. ICP Interference Check Sample (ICS) Analysis

The frequency of analysis was met.

The criteria for analysis were met.

V. Matrix Spike Analysis

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VI. Duplicate Sample Analysis

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Internal Standards

Raw data were not reviewed for this SDG.

IX. Furnace Atomic Absorption QC

Raw data were not reviewed for this SDG.

X. ICP Serial Dilution

ICP serial dilution analysis was performed by the laboratory. The analysis criteria were met with the following exceptions:

Diluted Sample	Analyte	%D (Limits)	Associated Samples	Flag	A or P
DUPE-5-2Q05L	Calcium Magnesium	13.5 (≤ 10) 10.2 (≤ 10)	All samples in SDG 05-2703	J (all detects) J (all detects)	A

XI. Sample Result Verification

Raw data were not reviewed for this SDG.

XII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XIII. Field Duplicates

Samples DUPE-6-2Q05 and MW-25-2 were identified as field duplicates. No metals were detected in any of the samples with the following exceptions:

Analyte	Concentration (ug/L)		RPD
	DUPE-6-2Q05	MW-25-2	
Calcium	13000	12500	4
Chromium	3.5	3.2	9
Iron	1500	1430	5
Lead	0.053	0.060	12
Magnesium	18700	18700	0
Potassium	3500	3460	1
Sodium	87600	88100	1

XIV. Field Blanks

Sample EB-11-5/12/05 was identified as an equipment blank. No metal contaminants were found in this blank with the following exceptions:

Equipment Blank ID	Analyte	Concentration (ug/L)
EB-11-5/12/05	Chromium	2.2
	Lead	0.045
	Magnesium	26.2
	Potassium	110

NASA JPL
Metals - Data Qualification Summary - SDG 05-2703

SDG	Sample	Analyte	Flag	A or P	Reason
05-2703	DUPE-6-2Q05 EB-11-5/12/05 MW-25-1 MW-25-2 MW-25-3 MW-25-4 MW-25-5	Calcium Magnesium	J (all detects) J (all detects)	A	ICP serial dilution (%D)

NASA JPL
Metals - Laboratory Blank Data Qualification Summary - SDG 05-2703

SDG	Sample	Analyte	Modified Final Concentration	A or P
05-2703	EB-11-5/12/05	Magnesium Potassium	26.2U ug/L 110U ug/L	A

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 16, 2005
LDC Report Date: July 11, 2005
Matrix: Water
Parameters: Metals
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 05-2738

Sample Identification

DUPE-7-2Q05
EB-11-5/16/05
MW-11-1
MW-11-2
MW-11-3
MW-11-4
MW-11-5
MW-11-5MS
MW-11-5MSD
MW-11-5DUP

Introduction

This data review covers 10 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Methods 200.7, 200.8, and 200.9 for Metals. The metals analyzed were Arsenic, Calcium, Chromium, Iron, Lead, Magnesium, Potassium, and Sodium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blanks are summarized in Section III.

Field duplicates are summarized in Section XIII.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

An initial calibration was performed.

The frequency and analysis criteria of the initial calibration verification (ICV) and continuing calibration verification (CCV) were met.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the initial, continuing and preparation blanks with the following exceptions:

Method Blank ID	Analyte	Maximum Concentration	Associated Samples
PB (prep blank)	Potassium	113 ug/L	All samples in SDG 05-2738
ICB/CCB	Magnesium Potassium	16.337 ug/L 143.757 ug/L	All samples in SDG 05-2738

Data qualification by the initial, continuing and preparation blanks (ICB/CCB/PBs) was based on the maximum contaminant concentration in the ICB/CCB/PBs in the analysis of each analyte. The sample concentrations were either not detected or were significantly greater (>5X blank contaminants) than the concentrations found in the associated method blanks with the following exceptions:

Sample	Analyte	Reported Concentration	Modified Final Concentration
EB-11-5/16/05	Magnesium Potassium	14.8 ug/L 127 ug/L	14.8U ug/L 127U ug/L

IV. ICP Interference Check Sample (ICS) Analysis

The frequency of analysis was met.

The criteria for analysis were met.

V. Matrix Spike Analysis

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VI. Duplicate Sample Analysis

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

VIII. Internal Standards

Raw data were not reviewed for this SDG.

IX. Furnace Atomic Absorption QC

Raw data were not reviewed for this SDG.

X. ICP Serial Dilution

ICP serial dilution analysis was performed by the laboratory. The analysis criteria were met with the following exceptions:

Diluted Sample	Analyte	%D (Limits)	Associated Samples	Flag	A or P
MW-11-5L	Potassium	12.2 (≤ 10)	All samples in SDG 05-2738	J (all detects)	A

XI. Sample Result Verification

Raw data were not reviewed for this SDG.

XII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XIII. Field Duplicates

Samples DUPE-7-2Q05 and MW-11-3 were identified as field duplicates. No metals were detected in any of the samples with the following exceptions:

Analyte	Concentration (ug/L)		RPD
	DUPE-7-2Q05	MW-11-3	
Calcium	38800	38200	2
Chromium	8.1	7.6	6
Iron	849	815	4
Lead	0.055	0.11	67
Magnesium	14700	14500	1
Potassium	2370	2340	1
Sodium	26200	26200	0

XIV. Field Blanks

Sample EB-11-5/16/05 was identified as an equipment blank. No metal contaminants were found in this blank with the following exceptions:

Equipment Blank ID	Analyte	Concentration (ug/L)
EB-11-5/16/05	Chromium	2.5
	Lead	0.044
	Iron	10.2
	Magnesium	14.8
	Potassium	127

NASA JPL**Metals - Data Qualification Summary - SDG 05-2738**

SDG	Sample	Analyte	Flag	A or P	Reason
05-2738	DUPE-7-2Q05 EB-11-5/16/05 MW-11-1 MW-11-2 MW-11-3 MW-11-4 MW-11-5	Potassium	J (all detects)	A	ICP serial dilution (%D)

NASA JPL**Metals - Laboratory Blank Data Qualification Summary - SDG 05-2738**

SDG	Sample	Analyte	Modified Final Concentration	A or P
05-2738	EB-11-5/16/05	Magnesium Potassium	14.8U ug/L 127U ug/L	A

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 17, 2005
LDC Report Date: July 11, 2005
Matrix: Water
Parameters: Metals
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 05-2754

Sample Identification

EB-12-5/17/05
MW-23-1
MW-23-2
MW-23-3
MW-23-4
MW-23-5

Introduction

This data review covers 6 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Methods 200.7, 200.8, and 200.9 for Metals. The metals analyzed were Arsenic, Calcium, Chromium, Iron, Lead, Magnesium, Potassium, and Sodium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blanks are summarized in Section III.

Field duplicates are summarized in Section XIII.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

An initial calibration was performed.

The frequency and analysis criteria of the initial calibration verification (ICV) and continuing calibration verification (CCV) were met.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the initial, continuing and preparation blanks with the following exceptions:

Method Blank ID	Analyte	Maximum Concentration	Associated Samples
PB (prep blank)	Calcium Potassium	140 ug/L 45.1 ug/L	All samples in SDG 05-2754
ICB/CCB	Calcium Magnesium Potassium	385.894 ug/L 17.416 ug/L 68.048 ug/L	All samples in SDG 05-2754

Data qualification by the initial, continuing and preparation blanks (ICB/CCB/PBs) was based on the maximum contaminant concentration in the ICB/CCB/PBs in the analysis of each analyte. The sample concentrations were either not detected or were significantly greater (>5X blank contaminants) than the concentrations found in the associated method blanks with the following exceptions:

Sample	Analyte	Reported Concentration	Modified Final Concentration
EB-12-5/17/05	Magnesium Potassium	22.6 ug/L 49.2 ug/L	22.6U ug/L 49.2U ug/L

IV. ICP Interference Check Sample (ICS) Analysis

ICP interference check sample analysis was not required.

V. Matrix Spike Analysis

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VI. Duplicate Sample Analysis

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

VIII. Internal Standards

Raw data were not reviewed for this SDG.

IX. Furnace Atomic Absorption QC

Raw data were not reviewed for this SDG.

X. ICP Serial Dilution

ICP serial dilution analysis was performed by the laboratory. The analysis criteria were met with the following exceptions:

Diluted Sample	Analyte	%D (Limits)	Associated Samples	Flag	A or P
MW-12-5L	Potassium Chromium	14.2 (≤ 10) 13.0 (≤ 10)	All samples in SDG 05-2754	J (all detects) J (all detects)	A

XI. Sample Result Verification

Raw data were not reviewed for this SDG.

XII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XIII. Field Duplicates

No field duplicates were identified in this SDG.

XIV. Field Blanks

Sample EB-12-5/17/05 was identified as an equipment blank. No metal contaminants were found in this blank with the following exceptions:

Equipment Blank ID	Analyte	Concentration (ug/L)
EB-12-5/17/05	Chromium	0.98
	Lead	0.044
	Magnesium	22.6
	Potassium	49.2

NASA JPL
Metals - Data Qualification Summary - SDG 05-2754

SDG	Sample	Analyte	Flag	A or P	Reason
05-2754	EB-12-5/17/05 MW-23-1 MW-23-2 MW-23-3 MW-23-4 MW-23-5	Potassium Chromium	J (all detects) J (all detects)	A	ICP serial dilution (%D)

NASA JPL
Metals - Laboratory Blank Data Qualification Summary - SDG 05-2754

SDG	Sample	Analyte	Modified Final Concentration	A or P
05-2754	EB-12-5/17/05	Magnesium Potassium	22.6U ug/L 49.2U ug/L	A

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 18, 2005
LDC Report Date: July 12, 2005
Matrix: Water
Parameters: Metals
Validation Level: EPA Level III & IV
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2772

Sample Identification

EB-13-5/18/05
MW-12-1
MW-12-2
MW-12-3**
MW-12-4
MW-12-5
MW-12-5MS
MW-12-5MSD
MW-12-5DUP

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 9 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Methods 200.7, 200.8, and 200.9 for Metals. The metals analyzed were Arsenic, Calcium, Chromium, Iron, Lead, Magnesium, Potassium, and Sodium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blanks are summarized in Section III.

Field duplicates are summarized in Section XIII.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

An initial calibration was performed.

The frequency and analysis criteria of the initial calibration verification (ICV) and continuing calibration verification (CCV) were met.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the initial, continuing and preparation blanks with the following exceptions:

Method Blank ID	Analyte	Maximum Concentration	Associated Samples
PB (prep blank)	Calcium Potassium	140 ug/L 45.1 ug/L	All samples in SDG 05-2772
ICB/CCB	Calcium Magnesium Potassium	385.894 ug/L 17.416 ug/L 68.048 ug/L	All samples in SDG 05-2772

Data qualification by the initial, continuing and preparation blanks (ICB/CCB/PBs) was based on the maximum contaminant concentration in the ICB/CCB/PBs in the analysis of each analyte. The sample concentrations were either not detected or were significantly greater (>5X blank contaminants) than the concentrations found in the associated method blanks with the following exceptions:

Sample	Analyte	Reported Concentration	Modified Final Concentration
EB-13-5/18/05	Magnesium Potassium	18.6 ug/L 51.4 ug/L	18.6U ug/L 51.4U ug/L

IV. ICP Interference Check Sample (ICS) Analysis

ICP interference check sample analysis was not required.

V. Matrix Spike Analysis

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VI. Duplicate Sample Analysis

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Internal Standards

All internal standard percent recoveries (%R) were within QC limits for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

IX. Furnace Atomic Absorption QC

All graphite furnace atomic absorption QC were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for samples reviewed by Level III criteria.

X. ICP Serial Dilution

ICP serial dilution analysis was performed by the laboratory. The analysis criteria were met with the following exceptions:

Diluted Sample	Analyte	%D (Limits)	Associated Samples	Flag	A or P
MW-12-5L	Potassium Chromium	14.2 (≤ 10) 13.0 (≤ 10)	All samples in SDG 05-2772	J (all detects) J (all detects)	A

XI. Sample Result Verification

All sample result verifications were acceptable for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XIII. Field Duplicates

No field duplicates were identified in this SDG.

XIV. Field Blanks

Sample EB-13-5/18/05 was identified as an equipment blank. No metal contaminants were found in this blank with the following exceptions:

Equipment Blank ID	Analyte	Concentration (ug/L)
EB-13-5/18/05	Chromium	0.30
	Lead	0.030
	Magnesium	18.6
	Potassium	51.4

NASA JPL**Metals - Data Qualification Summary - SDG 05-2772**

SDG	Sample	Analyte	Flag	A or P	Reason
05-2772	EB-13-5/18/05 MW-12-1 MW-12-2 MW-12-3** MW-12-4 MW-12-5	Potassium Chromium	J (all detects) J (all detects)	A	ICP serial dilution (%D)

NASA JPL**Metals - Laboratory Blank Data Qualification Summary - SDG 05-2772**

SDG	Sample	Analyte	Modified Final Concentration	A or P
05-2772	EB-13-5/18/05	Magnesium Potassium	18.6U ug/L 51.4U ug/L	A

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 19, 2005
LDC Report Date: July 12, 2005
Matrix: Water
Parameters: Metals
Validation Level: EPA Level III & IV
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 05-2788

Sample Identification

EB-14-5/19/05
MW-4-1**
MW-4-2
MW-4-3
MW-4-4
MW-4-5

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 6 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Methods 200.7, 200.8, and 200.9 for Metals. The metals analyzed were Arsenic, Calcium, Chromium, Iron, Lead, Magnesium, Potassium, and Sodium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (October 2004) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blanks are summarized in Section III.

Field duplicates are summarized in Section XIII.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

An initial calibration was performed.

The frequency and analysis criteria of the initial calibration verification (ICV) and continuing calibration verification (CCV) were met.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the initial, continuing and preparation blanks with the following exceptions:

Method Blank ID	Analyte	Maximum Concentration	Associated Samples
PB (prep blank)	Calcium Potassium	140 ug/L 45.1 ug/L	All samples in SDG 05-2788
ICB/CCB	Calcium Magnesium Potassium	385.894 ug/L 17.416 ug/L 68.048 ug/L	All samples in SDG 05-2788

Data qualification by the initial, continuing and preparation blanks (ICB/CCB/PBs) was based on the maximum contaminant concentration in the ICB/CCB/PBs in the analysis of each analyte. The sample concentrations were either not detected or were significantly greater (>5X blank contaminants) than the concentrations found in the associated method blanks with the following exceptions:

Sample	Analyte	Reported Concentration	Modified Final Concentration
EB-14-5/19/05	Calcium Magnesium Potassium	320 ug/L 15.6 ug/L 58.3 ug/L	320U ug/L 15.6U ug/L 58.3U ug/L

IV. ICP Interference Check Sample (ICS) Analysis

ICP interference check sample analysis was not required.

V. Matrix Spike Analysis

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VI. Duplicate Sample Analysis

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Internal Standards

All internal standard percent recoveries (%R) were within QC limits for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

IX. Furnace Atomic Absorption QC

All graphite furnace atomic absorption QC were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for samples reviewed by Level III criteria.

X. ICP Serial Dilution

ICP serial dilution analysis was performed by the laboratory. The analysis criteria were met with the following exceptions:

Diluted Sample	Analyte	%D (Limits)	Associated Samples	Flag	A or P
MW-12-5L	Potassium Chromium	14.2 (≤ 10) 13.0 (≤ 10)	All samples in SDG 05-2788	J (all detects) J (all detects)	A

XI. Sample Result Verification

All sample result verifications were acceptable for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XIII. Field Duplicates

No field duplicates were identified in this SDG.

XIV. Field Blanks

Sample EB-14-5/19/05 was identified as an equipment blank. No metal contaminants were found in this blank with the following exceptions:

Equipment Blank ID	Analyte	Concentration (ug/L)
EB-14-5/19/05	Chromium	2.6
	Lead	0.026
	Calcium	320
	Iron	7.1
	Magnesium	15.6
	Potassium	58.3

NASA JPL**Metals - Data Qualification Summary - SDG 05-2788**

SDG	Sample	Analyte	Flag	A or P	Reason
05-2788	EB-14-5/19/05 MW-4-1** MW-4-2 MW-4-3 MW-4-4 MW-4-5	Potassium Chromium	J (all detects) J (all detects)	A	ICP serial dilution (%D)

NASA JPL**Metals - Laboratory Blank Data Qualification Summary - SDG 05-2788**

SDG	Sample	Analyte	Modified Final Concentration	A or P
05-2788	EB-14-5/19/05	Calcium Magnesium Potassium	320U ug/L 15.6U ug/L 58.3U ug/L	A

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 20, 2005
LDC Report Date: July 11, 2005
Matrix: Water
Parameters: Metals
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 05-2812

Sample Identification

DUPE-8-2Q05
MW-5
MW-6
MW-16
MW-5MS
MW-5MSD
MW-5DUP

Introduction

This data review covers 7 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Methods 200.7, 200.8, and 200.9 for Metals. The metals analyzed were Arsenic, Calcium, Chromium, Iron, Lead, Magnesium, Potassium, and Sodium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blanks are summarized in Section III.

Field duplicates are summarized in Section XIII.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

An initial calibration was performed.

The frequency and analysis criteria of the initial calibration verification (ICV) and continuing calibration verification (CCV) were met.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the initial, continuing and preparation blanks with the following exceptions:

Method Blank ID	Analyte	Maximum Concentration	Associated Samples
PB (prep blank)	Potassium	60.9 ug/L	All samples in SDG 05-2812
ICB/CCB	Magnesium Potassium	7.898 ug/L 67.422 ug/L	All samples in SDG 05-2812

Data qualification by the initial, continuing and preparation blanks (ICB/CCB/PBs) was based on the maximum contaminant concentration in the ICB/CCB/PBs in the analysis of each analyte. The sample concentrations were either not detected or were significantly greater (>5X blank contaminants) than the concentrations found in the associated method blanks.

IV. ICP Interference Check Sample (ICS) Analysis

The frequency of analysis was met.

The criteria for analysis were met.

V. Matrix Spike Analysis

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VI. Duplicate Sample Analysis

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Internal Standards

Raw data were not reviewed for this SDG.

IX. Furnace Atomic Absorption QC

Raw data were not reviewed for this SDG.

X. ICP Serial Dilution

ICP serial dilution analysis was performed by the laboratory. The analysis criteria were met.

XI. Sample Result Verification

Raw data were not reviewed for this SDG.

XII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XIII. Field Duplicates

Samples DUPE-8-2Q05 and MW-6 were identified as field duplicates. No metals were detected in any of the samples with the following exceptions:

Analyte	Concentration (ug/L)		RPD
	DUPE-8-2Q05	MW-6	
Arsenic	2.0	1.9	5
Calcium	138000	139000	1
Chromium	13.0	13.6	5
Iron	368	411	11

Analyte	Concentration (ug/L)		RPD
	DUPE-8-2Q05	MW-6	
Lead	0.034	0.030	13
Magnesium	45700	45700	0
Potassium	2770	2770	0
Sodium	34300	34600	1

XIV. Field Blanks

No field blanks were identified in this SDG.

NASA JPL

Metals - Data Qualification Summary - SDG 05-2812

No Sample Data Qualified in this SDG

NASA JPL

Metals - Laboratory Blank Data Qualification Summary - SDG 05-2812

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 23, 2005
LDC Report Date: July 11, 2005
Matrix: Water
Parameters: Metals
Validation Level: EPA Level IV
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2820
Sample Identification
MW-13

Introduction

This data review covers one water sample listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Methods 200.7, 200.8, and 200.9 for Metals. The metals analyzed were Arsenic, Calcium, Chromium, Iron, Lead, Magnesium, Potassium, and Sodium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blanks are summarized in Section III.

Field duplicates are summarized in Section XIII.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

An initial calibration was performed.

The frequency and analysis criteria of the initial calibration verification (ICV) and continuing calibration verification (CCV) were met.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the initial, continuing and preparation blanks with the following exceptions:

Method Blank ID	Analyte	Maximum Concentration	Associated Samples
PB (prep blank)	Potassium	60.9 ug/L	All samples in SDG 05-2820
ICB/CCB	Magnesium Potassium	7.898 ug/L 67.422 ug/L	All samples in SDG 05-2820

Data qualification by the initial, continuing and preparation blanks (ICB/CCB/PBs) was based on the maximum contaminant concentration in the ICB/CCB/PBs in the analysis of each analyte. The sample concentrations were either not detected or were significantly greater (>5X blank contaminants) than the concentrations found in the associated method blanks.

IV. ICP Interference Check Sample (ICS) Analysis

The frequency of analysis was met.

The criteria for analysis were met.

V. Matrix Spike Analysis

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VI. Duplicate Sample Analysis

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Internal Standards

All internal standard percent recoveries (%R) were within QC limits.

IX. Furnace Atomic Absorption QC

All graphite furnace atomic absorption QC were within validation criteria.

X. ICP Serial Dilution

ICP serial dilution analysis was performed by the laboratory. The analysis criteria were met.

XI. Sample Result Verification

All sample result verifications were acceptable.

XII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XIII. Field Duplicates

No field duplicates were identified in this SDG.

XIV. Field Blanks

No field blanks were identified in this SDG.

NASA JPL

Metals - Data Qualification Summary - SDG 05-2820

No Sample Data Qualified in this SDG

NASA JPL

Metals - Laboratory Blank Data Qualification Summary - SDG 05-2820

No Sample Data Qualified in this SDG

Applied P & CH Laboratories
Metal Analysis Results

Client Name:	Battelle - Columbus Operations	Project No:	G486090	Collection Date:	05/23/2005
Project ID:	JPL GW Mon 2Q05	Service ID:	052820	Collected by:	
		Lab Sample ID:	05-2820-1	Received Date:	05/23/2005
Sample ID:	MW-13	Sample Matrix	Water	Moisture %:	-
Sample Type:	Field Sample				

Element Name	CAS No	Unit	RL	Result	C	M	Q	Batch	D-Date	A-Date	DF	Method
CHROMIUM		µg/L	1	25.7		MS		05M1520Q	05/24/05	05/25/05	1	200.8
LEAD		µg/L	1	0.039	B	MS		05M1520Q	05/24/05	05/25/05	1	200.8
ARSENIC	7440-38-2	µg/L	5	1.3	B	F		05M1547E	06/01/05	06/01/05	1	200.9
CALCIUM	7440-70-2	µg/L	200	64500		P		05M1512L	05/24/05	05/24/05	1	200.7
IRON	7439-89-6	µg/L	50	61.6		P		05M1512L	05/24/05	05/24/05	1	200.7
MAGNESIUM	7439-95-4	µg/L	100	22400		P		05M1512L	05/24/05	05/24/05	1	200.7
POTASSIUM	7440-09-7	µg/L	400	2880		P		05M1512L	05/24/05	05/24/05	1	200.7
SODIUM	7440-23-5	µg/L	2000	30600		P		05M1512L	05/24/05	05/24/05	1	200.7

Note: RL: PQL (EQL) or CRDL D-Date: Digestion Date; A-Date: Analysis Date; DF: Dilution Factor
C Qualifier: U - Not Detected or less than IDL B - Less than RL (PQL, EQL or CRDL), but greater than IDL.
Q Qualifier: N - Spike recovery out of control * - Duplicate analysis out of control
 W - Post digestion spike for GFAA out of control E - Serial dilution difference out of control
M Qualifier: MS - ICPMS P - ICP A - FLAA F - GFAA CV - Cold Vapor

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 24, 2005
LDC Report Date: July 12, 2005
Matrix: Water
Parameters: Metals
Validation Level: EPA Level III & IV
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 05-2832

Sample Identification

DUPE-9-2Q05
MW-7
MW-8**
MW-10
MW-7MS
MW-7MSD
MW-7DUP
MW-8MS
MW-8MSD

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 9 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Methods 200.7, 200.8, and 200.9 for Metals. The metals analyzed were Arsenic, Calcium, Chromium, Iron, Lead, Magnesium, Potassium, and Sodium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (October 2004) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blanks are summarized in Section III.

Field duplicates are summarized in Section XIII.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

An initial calibration was performed.

The frequency and analysis criteria of the initial calibration verification (ICV) and continuing calibration verification (CCV) were met.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the initial, continuing and preparation blanks with the following exceptions:

Method Blank ID	Analyte	Maximum Concentration	Associated Samples
PB (prep blank)	Potassium	57.5 ug/L	All samples in SDG 05-2832
ICB/CCB	Calcium Chromium Potassium	151.855 ug/L 0.088 ug/L 60.446 ug/L	All samples in SDG 05-2832

Data qualification by the initial, continuing and preparation blanks (ICB/CCB/PBs) was based on the maximum contaminant concentration in the ICB/CCB/PBs in the analysis of each analyte. The sample concentrations were either not detected or were significantly greater (>5X blank contaminants) than the concentrations found in the associated method blanks.

IV. ICP Interference Check Sample (ICS) Analysis

ICP interference check sample analysis was not required.

V. Matrix Spike Analysis

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VI. Duplicate Sample Analysis

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Internal Standards

All internal standard percent recoveries (%R) were within QC limits for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

IX. Furnace Atomic Absorption QC

All graphite furnace atomic absorption QC were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for samples reviewed by Level III criteria.

X. ICP Serial Dilution

ICP serial dilution analysis was performed by the laboratory. The analysis criteria were met with the following exceptions:

Diluted Sample	Analyte	%D (Limits)	Associated Samples	Flag	A or P
MW-7L	Iron	17.0 (≤ 10)	All samples in SDG 05-2832	J (all detects)	A

XI. Sample Result Verification

All sample result verifications were acceptable for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XIII. Field Duplicates

Samples DUPE-9-2Q05 and MW-10 were identified as field duplicates. No metals were detected in any of the samples with the following exceptions:

Compound	Concentration (ug/L)		RPD
	DUPE-9-2Q05	MW-10	
Calcium	104000	102000	2
Chromium	22.2	21.1	5
Iron	39.7	34.8	13
Lead	0.025	0.031	21
Magnesium	35700	35300	1
Potassium	3620	3560	2
Sodium	26500	25900	2

XIV. Field Blanks

No field blanks were identified in this SDG.

NASA JPL
Metals - Data Qualification Summary - SDG 05-2832

SDG	Sample	Analyte	Flag	A or P	Reason
05-2832	DUPE-9-2Q05 MW-7 MW-8** MW-10	Iron	J (all detects)	A	ICP serial dilution (%D)

NASA JPL
Metals - Laboratory Blank Data Qualification Summary - SDG 05-2832

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 25, 2005
LDC Report Date: July 13, 2005
Matrix: Water
Parameters: Metals
Validation Level: EPA Level III & IV
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2848

Sample Identification

EB-19-5/25/05
MW-26-1
MW-26-2**
SB-1-2Q05

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 4 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Methods 200.7, 200.8, and 200.9 for Metals. The metals analyzed were Arsenic, Calcium, Chromium, Iron, Lead, Magnesium, Potassium, and Sodium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (October 2004) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blanks are summarized in Section III.

Field duplicates are summarized in Section XIII.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

An initial calibration was performed.

The frequency and analysis criteria of the initial calibration verification (ICV) and continuing calibration verification (CCV) were met.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the initial, continuing and preparation blanks with the following exceptions:

Method Blank ID	Analyte	Maximum Concentration	Associated Samples
PB (prep blank)	Potassium	57.5 ug/L	All samples in SDG 05-2848
ICB/CCB	Calcium Chromium Potassium	151.855 ug/L 0.088 ug/L 60.446 ug/L	All samples in SDG 05-2848

Data qualification by the initial, continuing and preparation blanks (ICB/CCB/PBs) was based on the maximum contaminant concentration in the ICB/CCB/PBs in the analysis of each analyte. The sample concentrations were either not detected or were significantly greater (>5X blank contaminants) than the concentrations found in the associated method blanks with the following exceptions:

Sample	Analyte	Reported Concentration	Modified Final Concentration
EB-19-5/25/05	Potassium	68.2 ug/L	68.2U ug/L
SB-1-2Q05	Potassium	78.8 ug/L	78.8U ug/L

IV. ICP Interference Check Sample (ICS) Analysis

ICP interference check sample analysis was not required.

V. Matrix Spike Analysis

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VI. Duplicate Sample Analysis

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Internal Standards

All internal standard percent recoveries (%R) were within QC limits for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

IX. Furnace Atomic Absorption QC

All graphite furnace atomic absorption QC were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for samples reviewed by Level III criteria.

X. ICP Serial Dilution

ICP serial dilution analysis was performed by the laboratory. The analysis criteria were met with the following exceptions:

Diluted Sample	Analyte	%D (Limits)	Associated Samples	Flag	A or P
MW-7L	Iron	17.0 (≤ 10)	All samples in SDG 05-2848	J (all detects)	A

XI. Sample Result Verification

All sample result verifications were acceptable for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XIII. Field Duplicates

No field duplicates were identified in this SDG.

XIV. Field Blanks

Sample EB-19-5/25/05 was identified as an equipment blank. No metal contaminants were found in this blank with the following exceptions:

Equipment Blank ID	Analyte	Concentration (ug/L)
EB-19-5/25/05	Chromium	0.61
	Arsenic	1.9
	Iron	17.1
	Magnesium	12.6
	Potassium	68.2
	Sodium	459

Sample SB-1-2Q05 was identified as a source blank. No metal contaminants were found in this blank with the following exceptions:

Source Blank ID	Analyte	Concentration (ug/L)
SB-1-2Q05	Chromium	1.5
	Iron	22.6
	Magnesium	16.2
	Potassium	78.8

NASA JPL
Metals - Data Qualification Summary - SDG 05-2848

SDG	Sample	Analyte	Flag	A or P	Reason
05-2848	EB-19-5/25/05 MW-26-1 MW-26-2** SB-1-2Q05	Iron	J (all detects)	A	ICP serial dilution (%D)

NASA JPL
Metals - Laboratory Blank Data Qualification Summary - SDG 05-2848

SDG	Sample	Analyte	Modified Final Concentration	A or P
05-2848	EB-19-5/25/05	Potassium	68.2U ug/L	A
05-2848	SB-1-2Q05	Potassium	78.8U ug/L	A

NASA JPL
Data Validation Reports

Perchlorate

LDC

Laboratory Data Consultants, Inc.
Data Validation Report

Project/Site Name: NASA JPL
Collection Date: April 26, 2005
LDC Report Date: June 24, 2005
Matrix: Water
Parameters: Perchlorate
Validation Level: EPA Level III & IV
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2456

Sample Identification

EB-1-4/26/05
MW-21-1
MW-21-2**
MW-21-3
MW-21-4
MW-21-5
MW-21-2MS
MW-21-2MSD

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 8 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 314.0 for Perchlorate.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified a P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No perchlorate contaminants were found in the method blanks.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

The laboratory has indicated that there were no duplicate (DUP) analyses specified for the samples in this SDG, and therefore duplicate analyses were not performed for this SDG.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification

All sample result verifications were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

IX. Field Duplicates

No field duplicates were identified in this SDG.

X. Field Blanks

Sample EB-1-4/26/05 was identified as an equipment blank. No perchlorate contaminants were found in this blank.

NASA JPL

Perchlorate - Data Qualification Summary - SDG 05-2456

No Sample Data Qualified in this SDG

NASA JPL

Perchlorate - Laboratory Blank Data Qualification Summary - SDG 05-2456

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: April 27, 2005
LDC Report Date: June 24, 2005
Matrix: Water
Parameters: Perchlorate
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2486

Sample Identification

EB-2-4/27/05
MW-19-1
MW-19-2
MW-19-3
MW-19-4
MW-19-5
MW-19-2MS
MW-19-2MSD

Introduction

This data review covers 8 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 314.0 for Perchlorate.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified a P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No perchlorate contaminants were found in the method blanks.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

The laboratory has indicated that there were no duplicate (DUP) analyses specified for the samples in this SDG, and therefore duplicate analyses were not performed for this SDG.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification

Raw data were not reviewed for this SDG.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

IX. Field Duplicates

No field duplicates were identified in this SDG.

X. Field Blanks

Sample EB-2-4/27/05 was identified as an equipment blank. No perchlorate contaminants were found in this blank.

NASA JPL

Perchlorate - Data Qualification Summary - SDG 05-2486

No Sample Data Qualified in this SDG

NASA JPL

Perchlorate - Laboratory Blank Data Qualification Summary - SDG 05-2486

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: April 29, 2005
LDC Report Date: June 24, 2005
Matrix: Water
Parameters: Perchlorate
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2521

Sample Identification

EB-3-4/29/05
MW-20-1
MW-20-2
MW-20-3
MW-20-4
MW-20-5
MW-20-4MS
MW-20-4MSD

Introduction

This data review covers 8 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 314.0 for Perchlorate.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified a P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No perchlorate contaminants were found in the method blanks.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

The laboratory has indicated that there were no duplicate (DUP) analyses specified for the samples in this SDG, and therefore duplicate analyses were not performed for this SDG.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification

Raw data were not reviewed for this SDG.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

IX. Field Duplicates

No field duplicates were identified in this SDG.

X. Field Blanks

Sample EB-3-4/29/05 was identified as an equipment blank. No perchlorate contaminants were found in this blank.

NASA JPL

Perchlorate - Data Qualification Summary - SDG 05-2521

No Sample Data Qualified in this SDG

NASA JPL

Perchlorate - Laboratory Blank Data Qualification Summary - SDG 05-2521

No Sample Data Qualified in this SDG

Laboratory Data Consultants, Inc.
Data Validation Report

Project/Site Name: NASA JPL
Collection Date: May 2, 2005
LDC Report Date: June 24, 2005
Matrix: Water
Parameters: Perchlorate
Validation Level: EPA Level III & IV
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2546

Sample Identification

EB-4-5/2/05
MW-17-1
MW-17-2
MW-17-3
MW-17-4**
MW-17-5

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 6 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 314.0 for Perchlorate.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified a P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No perchlorate contaminants were found in the method blanks.

IV. Matrix Spike/Matrix Spike Duplicates

The laboratory has indicated that there were no matrix spike (MS) and matrix spike duplicate (MSD) analyses specified for the samples in this SDG, and therefore matrix spike and matrix spike duplicate analyses were not performed for this SDG.

V. Duplicates

The laboratory has indicated that there were no duplicate (DUP) analyses specified for the samples in this SDG, and therefore duplicate analyses were not performed for this SDG.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification

All sample result verifications were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

IX. Field Duplicates

No field duplicates were identified in this SDG.

X. Field Blanks

Sample EB-4-5/2/05 was identified as an equipment blank. No perchlorate contaminants were found in this blank.

NASA JPL

Perchlorate - Data Qualification Summary - SDG 05-2546

No Sample Data Qualified in this SDG

NASA JPL

Perchlorate - Laboratory Blank Data Qualification Summary - SDG 05-2546

No Sample Data Qualified in this SDG

Laboratory Data Consultants, Inc.
Data Validation Report

Project/Site Name: NASA JPL
Collection Date: May 3, 2005
LDC Report Date: June 24, 2005
Matrix: Water
Parameters: Perchlorate
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2556

Sample Identification

DUPE-1-2Q05
EB-5-5/3/05
MW-18-1
MW-18-2
MW-18-3
MW-18-4
MW-18-5

Introduction

This data review covers 7 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 314.0 for Perchlorate.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified a P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No perchlorate contaminants were found in the method blanks.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

The laboratory has indicated that there were no duplicate (DUP) analyses specified for the samples in this SDG, and therefore duplicate analyses were not performed for this SDG.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification

Raw data were not reviewed for this SDG.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

IX. Field Duplicates

Samples DUPE-1-2Q05 and MW-18-2 were identified as field duplicates. No perchlorate contaminants were detected in any of the samples.

X. Field Blanks

Sample EB-5-5/3/05 was identified as an equipment blank. No perchlorate contaminants were found in this blank.

NASA JPL

Perchlorate - Data Qualification Summary - SDG 05-2556

No Sample Data Qualified in this SDG

NASA JPL

Perchlorate - Laboratory Blank Data Qualification Summary - SDG 05-2556

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 6, 2005
LDC Report Date: June 24, 2005
Matrix: Water
Parameters: Perchlorate
Validation Level: EPA Level IV
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2624

Sample Identification

MW-1

Introduction

This data review covers one water sample listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 314.0 for Perchlorate.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified a P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No perchlorate contaminants were found in the method blanks.

IV. Matrix Spike/Matrix Spike Duplicates

The laboratory has indicated that there were no matrix spike (MS) and matrix spike duplicate (MSD) analyses specified for the samples in this SDG, and therefore matrix spike and matrix spike duplicate analyses were not performed for this SDG.

V. Duplicates

The laboratory has indicated that there were no duplicate (DUP) analyses specified for the samples in this SDG, and therefore duplicate analyses were not performed for this SDG.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification

All sample result verifications met validation criteria.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

IX. Field Duplicates

No field duplicates were identified in this SDG.

X. Field Blanks

No field blanks were identified in this SDG.

NASA JPL

Perchlorate - Data Qualification Summary - SDG 05-2624

No Sample Data Qualified in this SDG

NASA JPL

Perchlorate - Laboratory Blank Data Qualification Summary - SDG 05-2624

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 11, 2005
LDC Report Date: June 24, 2005
Matrix: Water
Parameters: Perchlorate
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2678

Sample Identification

EB-9-5/11/05
MW-24-1
MW-24-2
MW-24-3
MW-24-4
MW-24-5

Introduction

This data review covers 6 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 314.0 for Perchlorate.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified a P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No perchlorate contaminants were found in the method blanks.

IV. Matrix Spike/Matrix Spike Duplicates

The laboratory has indicated that there were no matrix spike (MS) and matrix spike duplicate (MSD) analyses specified for the samples in this SDG, and therefore matrix spike and matrix spike duplicate analyses were not performed for this SDG.

V. Duplicates

The laboratory has indicated that there were no duplicate (DUP) analyses specified for the samples in this SDG, and therefore duplicate analyses were not performed for this SDG.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification

Raw data were not reviewed for this SDG.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

IX. Field Duplicates

No field duplicates were identified in this SDG.

X. Field Blanks

Sample EB-9-5/11/05 was identified as an equipment blank. No perchlorate contaminants were found in this blank.

NASA JPL

Perchlorate - Data Qualification Summary - SDG 05-2678

No Sample Data Qualified in this SDG

NASA JPL

Perchlorate - Laboratory Blank Data Qualification Summary - SDG 05-2678

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 12, 2005
LDC Report Date: July 12, 2005
Matrix: Water
Parameters: Perchlorate
Validation Level: EPA Level III
Laboratory: Severn Trent Laboratories
Sample Delivery Group (SDG): G5E160164

Sample Identification

MW-25-5
MW-25-4
MW-25-3
MW-25-2
MW-25-1
MW-25-5RE
MW-25-4RE
MW-25-3RE
MW-25-2RE
MW-25-1RE
MW-25-5REMS
MW-25-5REMSD

Introduction

This data review covers 12 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA SW 846 Method 8321A for Perchlorate.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met with the following exceptions:

Sample	Compound	Total Days From Sample Collection Until Analysis	Required Holding Time (in Days) From Sample Collection Until Analysis	Flag	A or P
MW-25-5RE MW-25-4RE MW-25-3RE MW-25-2RE MW-25-1RE	Perchlorate	55	28	J (all detects) UJ (all non-detects)	A

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

Initial calibration of compounds was performed as required by the method.

A curve fit, based on the initial calibration, was established for quantitation. The coefficient of determination (r^2) was greater than or equal to 0.990 .

b. Calibration Verification

Calibration verification was performed at the required frequencies.

The percent differences (%D) of amounts in continuing standard mixtures were within the 30.0% QC limits.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No perchlorate contaminants were found in the method blanks.

IV. Accuracy and Precision Data

a. Surrogate Recovery

Surrogates were not required by the method.

b. Matrix Spike/(Matrix Spike) Duplicate

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

c. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

V. Target Compound Identification

Raw data were not reviewed for this SDG.

VI. Compound Quantitation and CRQLs

Raw data were not reviewed for this SDG.

VII. System Performance

Raw data were not reviewed for this SDG.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report.

IX. Field Duplicates

No field duplicates were identified in this SDG.

X. Field Blanks

No field blanks were identified in this SDG.

NASA JPL**Perchlorate - Data Qualification Summary - SDG G5E160164**

SDG	Sample	Compound	Flag	A or P	Reason
G5E160164	MW-25-5RE MW-25-4RE MW-25-3RE MW-25-2RE MW-25-1RE	Perchlorate	J (all detects) UJ (all non-detects)	A	Technical holding times

NASA JPL**Perchlorate - Laboratory Blank Data Qualification Summary - SDG G5E160164**

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 25, 2005
LDC Report Date: July 12, 2005
Matrix: Water
Parameters: Perchlorate
Validation Level: EPA Level III & IV
Laboratory: Severn Trent Laboratories
Sample Delivery Group (SDG): G5F020244

Sample Identification

MW-26-2**
MW-26-1
MW-26-2MS
MW-26-2MSD

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 4 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA SW 846 Method 8321A for Perchlorate.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A table summarizing all data qualification flags is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Samples indicated by a double asterisk on the front cover underwent EPA Level IV review. EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by EPA Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

Initial calibration of compounds was performed as required by the method.

A curve fit, based on the initial calibration, was established for quantitation. The coefficient of determination (r^2) was greater than or equal to 0.990 .

b. Calibration Verification

Calibration verification was performed at the required frequencies.

The percent differences (%D) of amounts in continuing standard mixtures were within the 30.0% QC limits.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No perchlorate was found in the method blanks.

IV. Accuracy and Precision Data

a. Surrogate Recovery

Surrogates were not required by the method.

b. Matrix Spike/(Matrix Spike) Duplicate

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

c. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

V. Target Compound Identification

All target compound identifications were within validation criteria for samples on which EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by EPA Level III criteria.

VI. Compound Quantitation and CRQLs

All compound quantitation and CRQLs were within validation criteria for samples on which EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by EPA Level III criteria.

VII. System Performance

The system performance was within validation criteria for samples on which EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by EPA Level III criteria.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report.

IX. Field Duplicates

No field duplicates were identified in this SDG.

X. Field Blanks

No field blanks were identified in this SDG.

NASA JPL

Perchlorate - Data Qualification Summary - SDG G5F020244

No Sample Data Qualified in this SDG

NASA JPL

Perchlorate - Laboratory Blank Data Qualification Summary - SDG G5F020244

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 12, 2005
LDC Report Date: June 24, 2005
Matrix: Water
Parameters: Perchlorate
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2702

Sample Identification

DUPE-6-2Q05
MW-25-1
MW-25-2
MW-25-3
MW-25-4
MW-25-5
EB-10-5/14/05

Introduction

This data review covers 7 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 314.0 for Perchlorate.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified a P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No perchlorate contaminants were found in the method blanks.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

The laboratory has indicated that there were no duplicate (DUP) analyses specified for the samples in this SDG, and therefore duplicate analyses were not performed for this SDG.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification

Raw data were not reviewed for this SDG.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

IX. Field Duplicates

Samples DUPE-6-2Q05 and MW-25-2 were identified as field duplicates. No perchlorate contaminants were detected in any of the samples with the following exceptions:

Analyte	Concentration (ug/L)		RPD
	DUPE-6-2Q05	MW-25-2	
Perchlorate	15.4	15.0	3

X. Field Blanks

Sample EB-10-5/14/05 was identified as an equipment blank. No perchlorate contaminants were found in this blank.

NASA JPL

Perchlorate - Data Qualification Summary - SDG 05-2702

No Sample Data Qualified in this SDG

NASA JPL

Perchlorate - Laboratory Blank Data Qualification Summary - SDG 05-2702

No Sample Data Qualified in this SDG

NASA JPL
Data Validation Reports

N-Nitrosodimethylamine

LDC

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 2, 2005
LDC Report Date: June 27, 2005
Matrix: Water
Parameters: N-Nitrosodimethylamine
Validation Level: EPA Level IV
Laboratory: Applied P & Ch Laboratory/Del Mar Analytical
Sample Delivery Group (SDG): 05-2547/IOE0213
Sample Identification
MW-17-4

Introduction

This data review covers one water sample listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 1625C for N-Nitrosodimethylamine.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for all compounds.

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% .

V. Blanks

Method blanks were reviewed for each matrix as applicable. No N-Nitrosodimethylamine contaminants were found in the method blanks.

VI. Surrogate Spikes

Surrogates were not required by the method.

VII. Matrix Spike/Matrix Spike Duplicates

The laboratory has indicated that there were no matrix spike (MS) and matrix spike duplicate (MSD) analyses specified for the samples in this SDG, and therefore matrix spike and matrix spike duplicate analyses were not performed for this SDG.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

All target compound identifications were within validation criteria.

XII. Compound Quantitation and CRQLs

All compound quantitation and CRQLs were within validation criteria.

XIII. Tentatively Identified Compounds (TICs)

Tentatively identified compounds were not reported by the laboratory.

XIV. System Performance

The system performance was acceptable.

XV. Overall Assessment of Data

Data flags have been summarized at the end of the report.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

XVII. Field Blanks

No field blanks were identified in this SDG.

NASA JPL

N-Nitrosodimethylamine - Data Qualification Summary - SDG 05-2547/IOE0213

No Sample Data Qualified in this SDG

NASA JPL

N-Nitrosodimethylamine - Laboratory Blank Data Qualification Summary - SDG 05-2547/IOE0213

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL

Collection Date: May 11, 2005

LDC Report Date: June 27, 2005

Matrix: Water

Parameters: N-Nitrosodimethylamine

Validation Level: EPA Level III

Laboratory: Applied P & Ch Laboratory/Del Mar Analytical

Sample Delivery Group (SDG): 05-2725/IOE1000

Sample Identification

MW-24-1

Introduction

This data review covers one water sample listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 1625C for N-Nitrosodimethylamine.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for all compounds.

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% .

V. Blanks

Method blanks were reviewed for each matrix as applicable. No N-Nitrosodimethylamine contaminants were found in the method blanks.

VI. Surrogate Spikes

Surrogates were not required by the method.

VII. Matrix Spike/Matrix Spike Duplicates

The laboratory has indicated that there were no matrix spike (MS) and matrix spike duplicate (MSD) analyses specified for the samples in this SDG, and therefore matrix spike and matrix spike duplicate analyses were not performed for this SDG.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits with the following exceptions:

LCS ID (Associated Samples)	Compound	LCS %R (Limits)	LCSD %R (Limits)	RPD (Limits)	Flag	A or P
5E18043-BS1/BSD1 (All samples in SDG 05-2725/IOE1000)	N-Nitrosodimethylamine	168 (70-130)	-	-	J (all detects)	P

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

Raw data were not reviewed for this SDG.

XII. Compound Quantitation and CRQLs

Raw data were not reviewed for this SDG.

XIII. Tentatively Identified Compounds (TICs)

Raw data were not reviewed for this SDG.

XIV. System Performance

Raw data were not reviewed for this SDG.

XV. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

XVII. Field Blanks

No field blanks were identified in this SDG.

NASA JPL

N-Nitrosodimethylamine - Data Qualification Summary - SDG 05-2725/IOE1000

SDG	Sample	Compound	Flag	A or P	Reason
05-2725/IOE1000	MW-24-1	N-Nitrosodimethylamine	J (all detects)	P	Laboratory control samples (%R)

NASA JPL

N-Nitrosodimethylamine - Laboratory Blank Data Qualification Summary - SDG 05-2725/IOE1000

No Sample Data Qualified in this SDG

NASA JPL
Data Validation Reports

1,2,3-Trichloropropane

LDC

Laboratory Data Consultants, Inc.
Data Validation Report

Project/Site Name: NASA JPL
Collection Date: May 12, 2005
LDC Report Date: July 11, 2005
Matrix: Water
Parameters: 1,2,3-Trichloropropane
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 05-2703

Sample Identification

DUPE-6-2Q05
EB-11-5/12/05
MW-25-1
MW-25-2
MW-25-3
MW-25-4
MW-25-5

Introduction

This data review covers 7 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 504.1 for 1,2,3-Trichloropropane.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

Initial calibration of compounds was performed for the primary (quantitation) column and confirmation column as required by this method.

The percent relative standard deviations (%RSD) were less than or equal to 20.0% for all compounds.

b. Calibration Verification

Calibration verification was performed at the required frequencies.

The percent differences (%D) of calibration factors in continuing standard mixtures were within the 20.0% QC limits.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No 1,2,3-Trichloropropane was found in the method blanks.

IV. Accuracy and Precision Data

a. Surrogate Recovery

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

b. Matrix Spike/(Matrix Spike) Duplicates

The laboratory has indicated that there were no matrix spike (MS) and matrix spike duplicate (MSD) analyses specified for the samples in this SDG, and therefore matrix spike and matrix spike duplicate analyses were not performed for this SDG.

c. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Target Compound Identification

Raw data were not reviewed for this SDG.

VI. Compound Quantitation and CRQLs

Raw data were not reviewed for this SDG.

VII. System Performance

Raw data were not reviewed for this SDG.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

IX. Field Duplicates

Samples DUPE-6-2Q05 and MW-25-2 were identified as field duplicates. No 1,2,3-Trichloropropane was detected in any of the samples.

X. Field Blanks

Sample EB-11-5/12/05 was identified as an equipment blank. No 1,2,3-Trichloropropane contaminants were found in this blank.

NASA JPL

1,2,3-Trichloropropane - Data Qualification Summary - SDG 05-2703

No Sample Data Qualified in this SDG

NASA JPL

1,2,3-Trichloropropane - Laboratory Blank Data Qualification Summary - SDG 05-2703

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 25, 2005
LDC Report Date: July 13, 2005
Matrix: Water
Parameters: 1,2,3-Trichloropropane
Validation Level: EPA Level III & IV
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2848

Sample Identification

MW-26-1
MW-26-2**

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 2 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 504.1 for 1,2,3-Trichloropropane.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Samples indicated by a double asterisk on the front cover underwent EPA Level IV review. EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by EPA Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

Initial calibration of compounds was performed for the primary (quantitation) column and confirmation column as required by this method.

The percent relative standard deviations (%RSD) were less than or equal to 20.0% for all compounds.

b. Calibration Verification

Calibration verification was performed at the required frequencies.

The percent differences (%D) of calibration factors in continuing standard mixtures were within the 20.0% QC limits.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No 1,2,3-Trichloropropane was found in the method blanks.

IV. Accuracy and Precision Data

a. Surrogate Recovery

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

b. Matrix Spike/(Matrix Spike) Duplicates

The laboratory has indicated that there were no matrix spike (MS) and matrix spike duplicate (MSD) analyses specified for the samples in this SDG, and therefore matrix spike and matrix spike duplicate analyses were not performed for this SDG.

c. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Target Compound Identification

All target compound identifications were within validation criteria for samples on which EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by EPA Level III criteria.

VI. Compound Quantitation and CRQLs

All compound quantitation and CRQLs were within validation criteria for samples on which EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by EPA Level III criteria.

VII. System Performance

The system performance was acceptable for samples on which EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by EPA Level III criteria.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

IX. Field Duplicates

No field duplicates were identified in this SDG.

X. Field Blanks

No field blanks were identified in this SDG.

NASA JPL

1,2,3-Trichloropropane - Data Qualification Summary - SDG 05-2848

No Sample Data Qualified in this SDG

NASA JPL

**1,2,3-Trichloropropane - Laboratory Blank Data Qualification Summary - SDG
05-2848**

No Sample Data Qualified in this SDG

NASA JPL
Data Validation Reports

1,4-Dioxane

INDC

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 2, 2005
LDC Report Date: July 11, 2005
Matrix: Water
Parameters: 1,4-Dioxane
Validation Level: EPA Level IV
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2545

Sample Identification

MW-17-4

Introduction

This data review covers one water sample listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA sw 846 Method 8270C using Selected Ion Monitoring (SIM) for 1,4-Dioxane.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for all compounds.

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% .

V. Blanks

Method blanks were reviewed for each matrix as applicable. No 1,4-dioxane was found in the method blanks.

VI. Surrogate Spikes

Surrogates were not required by the method.

VII. Matrix Spike/Matrix Spike Duplicates

The laboratory has indicated that there were no matrix spike (MS) and matrix spike duplicate (MSD) analyses specified for the samples in this SDG, and therefore matrix spike and matrix spike duplicate analyses were not performed for this SDG.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

All target compound identifications were within validation criteria.

XII. Compound Quantitation and CRQLs

All compound quantitation and CRQLs were within validation criteria.

XIII. Tentatively Identified Compounds (TICs)

Tentatively identified compounds were not reported by the laboratory.

XIV. System Performance

The system performance was acceptable.

XV. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

XVII. Field Blanks

No field blanks were identified in this SDG.

NASA JPL

1,4-Dioxane - Data Qualification Summary - SDG 05-2545

No Sample Data Qualified in this SDG

NASA JPL

1,4-Dioxane - Laboratory Blank Data Qualification Summary - SDG 05-2545

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 11, 2005
LDC Report Date: July 11, 2005
Matrix: Water
Parameters: 1,4-Dioxane
Validation Level: EPA Level IV
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2679

Sample Identification

MW-24-1

Introduction

This data review covers one water sample listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA sw 846 Method 8270C using Selected Ion Monitoring (SIM) for 1,4-Dioxane.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for all compounds.

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% .

V. Blanks

Method blanks were reviewed for each matrix as applicable. No 1,4-dioxane was found in the method blanks.

VI. Surrogate Spikes

Surrogates were not required by the method.

VII. Matrix Spike/Matrix Spike Duplicates

The laboratory has indicated that there were no matrix spike (MS) and matrix spike duplicate (MSD) analyses specified for the samples in this SDG, and therefore matrix spike and matrix spike duplicate analyses were not performed for this SDG.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

Raw data were not reviewed for this SDG.

XII. Compound Quantitation and CRQLs

Raw data were not reviewed for this SDG.

XIII. Tentatively Identified Compounds (TICs)

Raw data were not reviewed for this SDG.

XIV. System Performance

Raw data were not reviewed for this SDG.

XV. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

XVII. Field Blanks

No field blanks were identified in this SDG.

NASA JPL

1,4-Dioxane - Data Qualification Summary - SDG 05-2679

No Sample Data Qualified in this SDG

NASA JPL

1,4-Dioxane - Laboratory Blank Data Qualification Summary - SDG 05-2679

No Sample Data Qualified in this SDG

Laboratory Data Consultants, Inc.
Data Validation Report

Project/Site Name: NASA JPL
Collection Date: May 19, 2005
LDC Report Date: July 12, 2005
Matrix: Water
Parameters: 1,4-Dioxane
Validation Level: EPA Level IV
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2788
Sample Identification
MW-4-1

Introduction

This data review covers one water sample listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA sw 846 Method 8270C using Selected Ion Monitoring (SIM) for 1,4-Dioxane.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for all compounds.

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% .

V. Blanks

Method blanks were reviewed for each matrix as applicable. No 1,4-dioxane was found in the method blanks.

VI. Surrogate Spikes

Surrogates were not required by the method.

VII. Matrix Spike/Matrix Spike Duplicates

The laboratory has indicated that there were no matrix spike (MS) and matrix spike duplicate (MSD) analyses specified for the samples in this SDG, and therefore matrix spike and matrix spike duplicate analyses were not performed for this SDG.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

All target compound identifications were within validation criteria.

XII. Compound Quantitation and CRQLs

All compound quantitation and CRQLs were within validation criteria.

XIII. Tentatively Identified Compounds (TICs)

Tentatively identified compounds were not reported by the laboratory.

XIV. System Performance

The system performance was acceptable.

XV. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

XVII. Field Blanks

No field blanks were identified in this SDG.

NASA JPL

1,4-Dioxane - Data Qualification Summary - SDG 05-2788

No Sample Data Qualified in this SDG

NASA JPL

1,4-Dioxane - Laboratory Blank Data Qualification Summary - SDG 05-2788

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 20, 2005
LDC Report Date: July 12, 2005
Matrix: Water
Parameters: 1,4-Dioxane
Validation Level: EPA Level IV
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2812
Sample Identification
MW-16

Introduction

This data review covers one water sample listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA sw 846 Method 8270C using Selected Ion Monitoring (SIM) for 1,4-Dioxane.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for all compounds.

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% .

V. Blanks

Method blanks were reviewed for each matrix as applicable. No 1,4-dioxane was found in the method blanks.

VI. Surrogate Spikes

Surrogates were not required by the method.

VII. Matrix Spike/Matrix Spike Duplicates

The laboratory has indicated that there were no matrix spike (MS) and matrix spike duplicate (MSD) analyses specified for the samples in this SDG, and therefore matrix spike and matrix spike duplicate analyses were not performed for this SDG.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

Raw data were not reviewed for this SDG.

XII. Compound Quantitation and CRQLs

Raw data were not reviewed for this SDG.

XIII. Tentatively Identified Compounds (TICs)

Raw data were not reviewed for this SDG.

XIV. System Performance

Raw data were not reviewed for this SDG.

XV. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

XVII. Field Blanks

No field blanks were identified in this SDG.

NASA JPL

1,4-Dioxane - Data Qualification Summary - SDG 05-2812

No Sample Data Qualified in this SDG

NASA JPL

1,4-Dioxane - Laboratory Blank Data Qualification Summary - SDG 05-2812

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 23, 2005
LDC Report Date: July 12, 2005
Matrix: Water
Parameters: 1,4-Dioxane
Validation Level: EPA Level IV
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2820
Sample Identification
MW-13

Introduction

This data review covers one water sample listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA sw 846 Method 8270C using Selected Ion Monitoring (SIM) for 1,4-Dioxane.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for all compounds.

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% .

V. Blanks

Method blanks were reviewed for each matrix as applicable. No 1,4-dioxane was found in the method blanks.

VI. Surrogate Spikes

Surrogates were not required by the method.

VII. Matrix Spike/Matrix Spike Duplicates

The laboratory has indicated that there were no matrix spike (MS) and matrix spike duplicate (MSD) analyses specified for the samples in this SDG, and therefore matrix spike and matrix spike duplicate analyses were not performed for this SDG.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

All target compound identifications were within validation criteria.

XII. Compound Quantitation and CRQLs

All compound quantitation and CRQLs were within validation criteria.

XIII. Tentatively Identified Compounds (TICs)

Tentatively identified compounds were not reported by the laboratory.

XIV. System Performance

The system performance was acceptable.

XV. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

XVII. Field Blanks

No field blanks were identified in this SDG.

NASA JPL

1,4-Dioxane - Data Qualification Summary - SDG 05-2820

No Sample Data Qualified in this SDG

NASA JPL

1,4-Dioxane - Laboratory Blank Data Qualification Summary - SDG 05-2820

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 25, 2005
LDC Report Date: July 14, 2005
Matrix: Water
Parameters: 1,4-Dioxane
Validation Level: EPA Level III & IV
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2848

Sample Identification

MW-26-1
MW-26-2**

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 2 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA sw 846 Method 8270C using Selected Ion Monitoring (SIM) for 1,4-Dioxane.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Samples indicated by a double asterisk on the front cover underwent EPA Level IV review. EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by EPA Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for all compounds.

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% .

V. Blanks

Method blanks were reviewed for each matrix as applicable. No 1,4-dioxane was found in the method blanks.

VI. Surrogate Spikes

Surrogates were not required by the method.

VII. Matrix Spike/Matrix Spike Duplicates

The laboratory has indicated that there were no matrix spike (MS) and matrix spike duplicate (MSD) analyses specified for the samples in this SDG, and therefore matrix spike and matrix spike duplicate analyses were not performed for this SDG.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

All target compound identifications were within validation criteria for samples on which EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by EPA Level III criteria.

XII. Compound Quantitation and CRQLs

All compound quantitation and CRQLs were within validation criteria for samples on which EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by EPA Level III criteria.

XIII. Tentatively Identified Compounds (TICs)

Tentatively identified compounds were not reported by the laboratory.

XIV. System Performance

The system performance was acceptable for samples on which EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by EPA Level III criteria.

XV. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

XVII. Field Blanks

No field blanks were identified in this SDG.

NASA JPL

1,4-Dioxane - Data Qualification Summary - SDG 05-2848

No Sample Data Qualified in this SDG

NASA JPL

1,4-Dioxane - Laboratory Blank Data Qualification Summary - SDG 05-2848

No Sample Data Qualified in this SDG

NASA JPL
Data Validation Reports

Nitroaromatics & Nitramines

LDc

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 25, 2005
LDC Report Date: July 13, 2005
Matrix: Water
Parameters: Nitroaromatics and Nitramines
Validation Level: EPA Level III & IV
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2848

Sample Identification

MW-26-1
MW-26-2**

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 2 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA SW 846 Method 8330 for Nitroaromatics and Nitramines.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Samples indicated by a double asterisk on the front cover underwent EPA Level IV review. EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by EPA Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

Initial calibration of compounds was performed for the primary (quantitation) column and confirmation column as required by the method.

A curve fit, based on the initial calibration, was established for quantitation. The coefficient of determination (r^2) was greater than or equal to 0.990 .

b. Calibration Verification

Calibration verification was performed at the required frequencies. The percent differences (%D) of amounts in continuing standard mixtures were within the 15.0% QC limits.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No nitroaromatic or nitramine contaminants were found in the method blanks.

IV. Accuracy and Precision Data

a. Surrogate Recovery

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

b. Matrix Spike/(Matrix Spike) Duplicates

The laboratory has indicated that there were no matrix spike (MS) and matrix spike duplicate (MSD) analyses specified for the samples in this SDG, and therefore matrix spike and matrix spike duplicate analyses were not performed for this SDG.

c. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Target Compound Identification

All target compound identifications were within validation criteria for samples on which EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by EPA Level III criteria.

VI. Compound Quantitation and CRQLs

All compound quantitation and CRQLs were within validation criteria for samples on which EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by EPA Level III criteria.

VII. System Performance

The system performance was acceptable for samples on which EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by EPA Level III criteria.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

IX. Field Duplicates

No field duplicates were identified in this SDG.

X. Field Blanks

No field blanks were identified in this SDG.

NASA JPL

Nitroaromatics and Nitramines - Data Qualification Summary - SDG 05-2848

No Sample Data Qualified in this SDG

NASA JPL

Nitroaromatics and Nitramines - Laboratory Blank Data Qualification Summary - SDG 05-2848

No Sample Data Qualified in this SDG

NASA JPL
Data Validation Reports

Wet Chemistry

LDC

Laboratory Data Consultants, Inc.
Data Validation Report

Project/Site Name: NASA JPL
Collection Date: April 26, 2005
LDC Report Date: July 29, 2005
Matrix: Water
Parameters: Wet Chemistry
Validation Level: EPA Level III & IV
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2457

Sample Identification

EB-1-4/26/05
MW-21-1
MW-21-2**
MW-21-3
MW-21-4
MW-21-5
MW-21-2MS
MW-21-2MSD

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 8 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 160.1 for Total Dissolved Solids, EPA Method 300.0 for Chloride, Nitrate as Nitrogen, and Sulfate, Standard Method 2320B for Alkalinity, EPA SW 846 Method 7196 for Hexavalent Chromium, and EPA SW 846 Method 9040 for pH.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (October 2004) as there are no current guidelines for the methods stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification

All sample result verifications were acceptable for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

IX. Field Duplicates

No field duplicates were identified in this SDG.

X. Field Blanks

Sample EB-1-4/26/05 was identified as an equipment blank. No contaminant concentrations were found in this blank with the following exceptions:

Equipment Blank ID	Analyte	Concentration
EB-1-4/26/05	pH Total dissolved solids Chloride Nitrate as N Sulfate	7.12 units 7.0 mg/L 0.16 mg/L 0.097 mg/L 0.40 mg/L

NASA JPL

Wet Chemistry - Data Qualification Summary - SDG 05-2457

No Sample Data Qualified in this SDG

NASA JPL

Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 05-2457

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: April 27, 2005
LDC Report Date: July 29, 2005
Matrix: Water
Parameters: Wet Chemistry
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 05-2485

Sample Identification

EB-2-4/27/05
MW-19-1
MW-19-2
MW-19-3
MW-19-4
MW-19-5
EB-2-4/27/05DUP
MW-19-2MS
MW-19-2MSD

Introduction

This data review covers 9 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 160.1 for Total Dissolved Solids, EPA Method 300.0 for Chloride, Nitrate as Nitrogen, and Sulfate, Standard Method 2320B for Alkalinity, EPA SW 846 Method 7196 for Hexavalent Chromium, and EPA SW 846 Method 9040 for pH.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (October 2004) as there are no current guidelines for the methods stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification

Raw data were not reviewed for this SDG.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

IX. Field Duplicates

No field duplicates were identified in this SDG.

X. Field Blanks

Sample EB-2-4/27/05 was identified as an equipment blank. No contaminant concentrations were found in this blank with the following exceptions:

Equipment Blank ID	Analyte	Concentration
EB-2-4/27/05	pH Total dissolved solids Chloride Nitrate as N Sulfate	7.56 units 5.0 mg/L 0.12 mg/L 0.13 mg/L 0.44 mg/L

NASA JPL

Wet Chemistry - Data Qualification Summary - SDG 05-2485

No Sample Data Qualified in this SDG

NASA JPL

Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 05-2485

No Sample Data Qualified in this SDG

Laboratory Data Consultants, Inc.
Data Validation Report

Project/Site Name: NASA JPL
Collection Date: April 29, 2005
LDC Report Date: July 29, 2005
Matrix: Water
Parameters: Wet Chemistry
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2522

Sample Identification

EB-3-4/29/05
MW-20-1
MW-20-2
MW-20-3
MW-20-4
MW-20-5
MW-20-1MS
MW-20-1MSD
MW-20-4MS
MW-20-4MSD

Introduction

This data review covers 10 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 160.1 for Total Dissolved Solids, EPA Method 300.0 for Chloride, Nitrate as Nitrogen, and Sulfate, Standard Method 2320B for Alkalinity, EPA SW 846 Method 7196 for Hexavalent Chromium, and EPA SW 846 Method 9040 for pH.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (October 2004) as there are no current guidelines for the methods stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification

Raw data were not reviewed for this SDG.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

IX. Field Duplicates

No field duplicates were identified in this SDG.

X. Field Blanks

Sample EB-3-4/29/05 was identified as an equipment blank. No contaminant concentrations were found in this blank with the following exceptions:

Equipment Blank ID	Analyte	Concentration
EB-3-4/29/05	Total dissolved solids pH Chloride Nitrate as N Sulfate	6.0 mg/L 7.55 units 0.094 mg/L 0.038 mg/L 0.33 mg/L

NASA JPL

Wet Chemistry - Data Qualification Summary - SDG 05-2522

No Sample Data Qualified in this SDG

NASA JPL

Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 05-2522

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 2, 2005
LDC Report Date: July 7, 2005
Matrix: Water
Parameters: Wet Chemistry
Validation Level: EPA Level III & IV
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2545

Sample Identification

EB-4-5/2/05
MW-17-1
MW-17-2
MW-17-3
MW-17-4**
MW-17-5
EB-4-5/2/05DUP
MW-17-5MS
MW-17-5MSD

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 9 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 150.1 for pH, EPA Method 160.1 for Total Dissolved Solids, EPA Method 300.0 for Chloride, Nitrate as Nitrogen, and Sulfate, EPA Method 310.1 for Alkalinity, and EPA SW 846 Method 7196 for Hexavalent Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification

All sample result verifications were acceptable for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

IX. Field Duplicates

No field duplicates were identified in this SDG.

X. Field Blanks

Sample EB-4-5/2/05 was identified as an equipment blank. No contaminant concentrations were found in this blank with the following exceptions:

Equipment Blank ID	Analyte	Concentration
EB-4-5/2/05	pH	7.49 units
	Total dissolved solids	6.0 mg/L
	Chloride	0.24 mg/L
	Nitrate as N	0.051 mg/L
	Sulfate	0.36 mg/L

NASA JPL

Wet Chemistry - Data Qualification Summary - SDG 05-2545

No Sample Data Qualified in this SDG

NASA JPL

Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 05-2545

No Sample Data Qualified in this SDG

Laboratory Data Consultants, Inc.
Data Validation Report

Project/Site Name: NASA JPL
Collection Date: May 3, 2005
LDC Report Date: July 29, 2005
Matrix: Water
Parameters: Wet Chemistry
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2557

Sample Identification

DUPE-1-2Q05
EB-5-5/3/05
MW-18-1
MW-18-2
MW-18-3
MW-18-4
MW-18-5
DUPE-1-2Q05MS
DUPE-1-2Q05MSD
MW-18-5MS
MW-18-5MSD
MW-18-5DUP

Introduction

This data review covers 8 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 160.1 for Total Dissolved Solids, EPA Method 300.0 for Chloride, Nitrate as Nitrogen, and Sulfate, Standard Method 2320B for Alkalinity, EPA SW 846 Method 7196 for Hexavalent Chromium, and EPA SW 846 Method 9040 for pH.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (October 2004) as there are no current guidelines for the methods stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification

Raw data were not reviewed for this SDG.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

IX. Field Duplicates

Samples DUPE-1-2Q05 and MW-18-2 were identified as field duplicates. No contaminant concentrations were detected in any of the samples with the following exceptions:

Analyte	Concentration		RPD
	DUPE-1-2Q05	MW-18-2	
Alkalinity	188 mg/L	189 mg/L	1
Total dissolved solids	292 mg/L	278 mg/L	5
Chloride	14.1 mg/L	11.4 mg/L	21
Nitrate as N	0.98 mg/L	0.89 mg/L	10
Sulfate	34.5 mg/L	27.6 mg/L	22
pH	7.65 units	7.43 units	3

X. Field Blanks

Sample EB-5-5/3/05 was identified as an equipment blank. No contaminant concentrations were found in this blank with the following exceptions:

Equipment Blank ID	Analyte	Concentration
EB-5-5/3/05	pH Total dissolved solids Chloride Nitrate as N Sulfate	7.01 units 11.0 mg/L 0.15 mg/L 0.091 mg/L 0.38 mg/L

NASA JPL

Wet Chemistry - Data Qualification Summary - SDG 05-2557

No Sample Data Qualified in this SDG

NASA JPL

Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 05-2557

No Sample Data Qualified in this SDG

Laboratory Data Consultants, Inc.
Data Validation Report

Project/Site Name: NASA JPL
Collection Date: May 4, 2005
LDC Report Date: July 29, 2005
Matrix: Water
Parameters: Wet Chemistry
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2586

Sample Identification

EB-6-5/4/05
MW-3-1
MW-3-2
MW-3-3
MW-3-4
MW-3-5
EB-6-5/4/05DUP
MW-3-1MS
MW-3-1MSD
MW-3-3MS
MW-3-3MSD

Introduction

This data review covers 11 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 160.1 for Total Dissolved Solids, EPA Method 300.0 for Chloride, Nitrate as Nitrogen, and Sulfate, EPA Method 314.0 for Perchlorate, Standard Method 2320B for Alkalinity, EPA SW 846 Method 7196 for Hexavalent Chromium, and EPA SW 846 Method 9040 for pH.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (October 2004) as there are no current guidelines for the methods stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification

Raw data were not reviewed for this SDG.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

IX. Field Duplicates

No field duplicates were identified in this SDG.

X. Field Blanks

Sample EB-6-5/4/05 was identified as an equipment blank. No contaminant concentrations were found in this blank with the following exceptions:

Equipment Blank ID	Analyte	Concentration
EB-6-5/4/05	pH Total dissolved solids Chloride Nitrate as N Sulfate	6.82 units 5.0 mg/L 0.12 mg/L 0.095 mg/L 0.37 mg/L

NASA JPL

Wet Chemistry - Data Qualification Summary - SDG 05-2586

No Sample Data Qualified in this SDG

NASA JPL

Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 05-2586

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 6, 2005
LDC Report Date: July 29, 2005
Matrix: Water
Parameters: Wet Chemistry
Validation Level: EPA Level III & IV
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2625

Sample Identification

DUPE-2-2Q05
DUPE-3-2Q05
MW-1**
MW-9
MW-15**
MW-15MS
MW-15MSD

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 7 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 160.1 for Total Dissolved Solids, EPA Method 300.0 for Chloride, Nitrate as Nitrogen, and Sulfate, EPA Method 314.0 for Perchlorate, Standard Method 2320B for Alkalinity, EPA SW 846 Method 7196 for Hexavalent Chromium, and EPA SW 846 Method 9040 for pH.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (October 2004) as there are no current guidelines for the methods stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification

All sample result verifications were acceptable for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

IX. Field Duplicates

Samples DUPE-2-2Q05 and MW-1** and samples DUPE-3-2Q05 and MW-9 were identified as field duplicates. No contaminant concentrations were detected in any of the samples with the following exceptions:

Analyte	Concentration		RPD
	DUPE-2-2Q05	MW-1**	
Alkalinity	209 mg/L	205 mg/L	2
Total dissolved solids	341 mg/L	331 mg/L	3
Chloride	29.3 mg/L	28.0 mg/L	5
Nitrate as N	1.4 mg/L	1.4 mg/L	0
Sulfate	57.9 mg/L	55.4 mg/L	4
pH	7.17 units	7.30 units	2

Analyte	Concentration		RPD
	DUPE-3-2Q05	MW-9	
Alkalinity	163 mg/L	169 mg/L	4
Total dissolved solids	253 mg/L	270 mg/L	7
Chloride	16.2 mg/L	16.2 mg/L	0
Nitrate as N	0.97 mg/L	0.98 mg/L	1
Sulfate	38.9 mg/L	38.9 mg/L	0
pH	6.85 units	7.02 units	2

X. Field Blanks

No field blanks were identified in this SDG.

NASA JPL

Wet Chemistry - Data Qualification Summary - SDG 05-2625

No Sample Data Qualified in this SDG

NASA JPL

Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 05-2625

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 9, 2005
LDC Report Date: July 6, 2005
Matrix: Water
Parameters: Wet Chemistry
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 05-2652

Sample Identification

DUPE-4-2Q05
EB-7-5/9/05
MW-14-1
MW-14-2
MW-14-3
MW-14-4
MW-14-5
DUPE-4-2Q05MS
DUPE-4-2Q05MSD
DUPE-4-2Q05DUP
MW-14-5MS
MW-14-5MSD

Introduction

This data review covers 12 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 150.1 for pH, EPA Method 160.1 for Total Dissolved Solids, EPA Method 300.0 for Chloride, Nitrate as Nitrogen, and Sulfate, EPA Method 310.1 for Alkalinity, EPA Method 314.0 for Perchlorate, and EPA SW 846 Method 7196 for Hexavalent Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits with the following exceptions:

LCS ID (Associated Samples)	Analyte	LCS %R (Limits)	LCSD %R (Limits)	RPD (Limits)	Flag	A or P
LCS/LCSD (All samples in SDG 05-2652)	Perchlorate	83 (85-115)	-	-	J (all detects) UJ (all non-detects)	P

VII. Sample Result Verification

Raw data were not reviewed for this SDG.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

IX. Field Duplicates

Samples DUPE-4-2Q05 and MW-14-4 were identified as field duplicates. No contaminant concentrations were detected in any of the samples with the following exceptions:

Analyte	Concentration (mg/L)		RPD
	DUPE-4-2Q05	MW-14-4	
Alkalinity, total	162	220	30
Total dissolved solids	326	322	1
Chloride	40.6	41.7	3
Nitrate as N	10.6	10.8	2
Sulfate	34.7	35.1	1

Analyte	Concentration (units)		RPD
	DUPE-4-2Q05	MW-14-4	
pH	7.75	7.88	2

Analyte	Concentration (ug/L)		RPD
	DUPE-4-2Q05	MW-14-4	
Perchlorate	3.5	3.4	3

X. Field Blanks

Sample EB-7-5/9/05 was identified as an equipment blank. No contaminant concentrations were found in this blank with the following exceptions:

Equipment Blank ID	Analyte	Concentration
EB-7-5/9/05	pH Chloride Nitrate as N Sulfate	6.27 units 0.14 mg/L 0.097 mg/L 0.43 mg/L

NASA JPL**Wet Chemistry - Data Qualification Summary - SDG 05-2652**

SDG	Sample	Analyte	Flag	A or P	Reason
05-2652	DUPE-4-2Q05 EB-7-5/9/05 MW-14-1 MW-14-2 MW-14-3 MW-14-4 MW-14-5	Perchlorate	J (all detects) UJ (all non-detects)	P	Laboratory control samples (%R)

NASA JPL**Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 05-2652**

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 10, 2005
LDC Report Date: July 7, 2005
Matrix: Water
Parameters: Wet Chemistry
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2660

Sample Identification

DUPE-5-2Q05
EB-8-5/10/05
MW-22-1
MW-22-2
MW-22-3
MW-22-4
MW-22-5
DUPE-5-2Q05MS
DUPE-5-2Q05MSD
MW-22-5MS
MW-22-5MSD
MW-22-5DUP

Introduction

This data review covers 12 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 150.1 for pH, EPA Method 160.1 for Total Dissolved Solids, EPA Method 300.0 for Chloride, Nitrate as Nitrogen, and Sulfate, EPA Method 310.1 for Alkalinity, EPA Method 314.0 for Perchlorate, and EPA SW 846 Method 7196 for Hexavalent Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification

Raw data were not reviewed for this SDG.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

IX. Field Duplicates

Samples DUPE-5-2Q05 and MW-22-3 were identified as field duplicates. No contaminant concentrations were detected in any of the samples with the following exceptions:

Analyte	Concentration (mg/L)		RPD
	DUPE-5-2Q05	MW-22-3	
Alkalinity, total	162	169	4
Total dissolved solids	328	327	0
Chloride	36.5	36.3	1
Nitrate as N	7.0	6.8	3
Sulfate	42.2	40.4	4

Analyte	Concentration (units)		RPD
	DUPE-5-2Q05	MW-22-3	
pH	7.59	7.86	3

Analyte	Concentration (ug/L)		RPD
	DUPE-5-2Q05	MW-22-3	
Perchlorate	3.0	3.2	6

X. Field Blanks

Sample EB-8-5/10/05 was identified as an equipment blank. No contaminant concentrations were found in this blank with the following exceptions:

Equipment Blank ID	Analyte	Concentration
EB-8-5/10/05	pH Total dissolved solids Chloride Nitrate as N Sulfate	7.92 units 6.0 mg/L 0.33 mg/L 0.16 mg/L 1.0 mg/L

NASA JPL

Wet Chemistry - Data Qualification Summary - SDG 05-2660

No Sample Data Qualified in this SDG

NASA JPL

Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 05-2660

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 11, 2005
LDC Report Date: July 6, 2005
Matrix: Water
Parameters: Wet Chemistry
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2679

Sample Identification

EB-9-5/11/05
MW-24-1
MW-24-2
MW-24-3
MW-24-4
MW-24-5
EB-9-5/11/05MS
EB-9-5/11/05MSD
EB-9-5/11/05DUP
MW-24-1MS
MW-24-1MSD
MW-24-3MS
MW-24-3MSD

Introduction

This data review covers 13 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 150.1 for pH, EPA Method 160.1 for Total Dissolved Solids, EPA Method 300.0 for Chloride, Nitrate as Nitrogen, and Sulfate, EPA Method 310.1 for Alkalinity, and EPA SW 846 Method 7196 for Hexavalent Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification

Raw data were not reviewed for this SDG.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

IX. Field Duplicates

No field duplicates were identified in this SDG.

X. Field Blanks

Sample EB-9-5/11/05 was identified as an equipment blank. No contaminant concentrations were found in this blank with the following exceptions:

Equipment Blank ID	Analyte	Concentration
EB-9-5/11/05	pH Total dissolved solids Chloride Nitrate as N Sulfate	7.07 units 8.0 mg/L 0.14 mg/L 0.094 mg/L 0.37 mg/L

NASA JPL

Wet Chemistry - Data Qualification Summary - SDG 05-2679

No Sample Data Qualified in this SDG

NASA JPL

Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 05-2679

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 12, 2005
LDC Report Date: July 7, 2005
Matrix: Water
Parameters: Wet Chemistry
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 05-2703

Sample Identification

DUPE-6-2Q05
EB-11-5/12/05
MW-25-1
MW-25-2
MW-25-3
MW-25-4
MW-25-5
DUPE-6-2Q05MS
DUPE-6-2Q05MSD

Introduction

This data review covers 9 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 150.1 for pH, EPA Method 160.1 for Total Dissolved Solids, EPA Method 300.0 for Chloride, Nitrate as Nitrogen, and Sulfate, EPA Method 310.1 for Alkalinity, and EPA SW 846 Method 7196 for Hexavalent Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification

Raw data were not reviewed for this SDG.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

IX. Field Duplicates

Samples DUPE-6-2Q05 and MW-25-2 were identified as field duplicates. No contaminant concentrations were detected in any of the samples with the following exceptions:

Analyte	Concentration (mg/L)		RPD
	DUPE-6-2Q05	MW-25-2	
Alkalinity, total	139	135	3
Alkalinity, bicarbonate	99.1	92.4	7
Alkalinity, carbonate	40.2	42.8	6
Total dissolved solids	360	308	16
Chloride	40.8	34.5	17
Nitrate as N	3.5	3.2	9
Sulfate	88.3	82.6	7

Analyte	Concentration (units)		RPD
	DUPE-6-2Q05	MW-25-2	
pH	9.05	9.00	1

X. Field Blanks

Sample EB-11-5/12/05 was identified as an equipment blank. No contaminant concentrations were found in this blank with the following exceptions:

Equipment Blank ID	Analyte	Concentration
EB-11-5/12/05	pH Total dissolved solids Chloride Nitrate as N Sulfate	7.37 units 7.0 mg/L 0.50 mg/L 0.080 mg/L 0.36 mg/L

NASA JPL

Wet Chemistry - Data Qualification Summary - SDG 05-2703

No Sample Data Qualified in this SDG

NASA JPL

Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 05-2703

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 16, 2005
LDC Report Date: July 11, 2005
Matrix: Water
Parameters: Wet Chemistry
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 05-2738

Sample Identification

DUPE-7-2Q05
EB-11-5/16/05
MW-11-1
MW-11-2
MW-11-3
MW-11-4
MW-11-5
DUPE-7-2Q05MS
DUPE-7-2Q05MSD
DUPE-7-2Q05DUP

Introduction

This data review covers 10 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 150.1 for pH, EPA Method 160.1 for Total Dissolved Solids, EPA Method 300.0 for Chloride, Nitrate as Nitrogen, and Sulfate, EPA Method 310.1 for Alkalinity, EPA Method 314.0 for Perchlorate, and EPA SW 846 Method 7196 for Hexavalent Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification

Raw data were not reviewed for this SDG.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

IX. Field Duplicates

Samples DUPE-7-2Q05 and MW-11-3 were identified as field duplicates. No contaminant concentrations were detected in any of the samples with the following exceptions:

Analyte	Concentration (mg/L)		RPD
	DUPE-7-2Q05	MW-11-3	
Alkalinity	167	169	1
Total dissolved solids	214	237	10
Chloride	11.5	11.6	1
Nitrate as N	0.057	0.051	11
Sulfate	21.0	21.2	1

Analyte	Concentration (units)		RPD
	DUPE-7-2Q05	MW-11-3	
pH	7.87	7.85	0

X. Field Blanks

Sample EB-11-5/16/05 was identified as an equipment blank. No contaminant concentrations were found in this blank with the following exceptions:

Equipment Blank ID	Analyte	Concentration
EB-11-5/16/05	pH Total dissolved solids Chloride Nitrate as N Sulfate	6.49 units 14.0 mg/L 0.13 mg/L 0.076 mg/L 0.36 mg/L

NASA JPL

Wet Chemistry - Data Qualification Summary - SDG 05-2738

No Sample Data Qualified in this SDG

NASA JPL

Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 05-2738

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 17, 2005
LDC Report Date: July 11, 2005
Matrix: Water
Parameters: Wet Chemistry
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 05-2754

Sample Identification

EB-12-5/17/05
MW-23-1
MW-23-2
MW-23-3
MW-23-4
MW-23-5
EB-12-5/17/05DUP
MW-23-1MS
MW-23-1MSD
MW-23-1DUP

Introduction

This data review covers 10 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 150.1 for pH, EPA Method 160.1 for Total Dissolved Solids, EPA Method 300.0 for Chloride, Nitrate as Nitrogen, and Sulfate, EPA Method 310.1 for Alkalinity, EPA Method 314.0 for Perchlorate, and EPA SW 846 Method 7196 for Hexavalent Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification

Raw data were not reviewed for this SDG.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

IX. Field Duplicates

No field duplicates were identified in this SDG.

X. Field Blanks

Sample EB-12-5/17/05 was identified as an equipment blank. No contaminant concentrations were found in this blank with the following exceptions:

Equipment Blank ID	Analyte	Concentration
EB-12-5/17/05	pH Total dissolved solids Chloride Nitrate as N	6.74 units 7.0 mg/L 0.11 mg/L 0.030 mg/L

NASA JPL

Wet Chemistry - Data Qualification Summary - SDG 05-2754

No Sample Data Qualified in this SDG

NASA JPL

Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 05-2754

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 18, 2005
LDC Report Date: July 12, 2005
Matrix: Water
Parameters: Wet Chemistry
Validation Level: EPA Level III & IV
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 05-2772

Sample Identification

EB-13-5/18/05
MW-12-1
MW-12-2
MW-12-3**
MW-12-4
MW-12-5
MW-12-2DUP
MW-12-5MS
MW-12-5MSD

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 9 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 150.1 for pH, EPA Method 160.1 for Total Dissolved Solids, EPA Method 300.0 for Chloride, Nitrate as Nitrogen, and Sulfate, EPA Method 310.1 for Alkalinity, EPA Method 314.0 for Perchlorate, and EPA SW 846 Method 7196 for Hexavalent Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (October 2004) as there are no current guidelines for the methods stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification

All sample result verifications were acceptable for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

IX. Field Duplicates

No field duplicates were identified in this SDG.

X. Field Blanks

Sample EB-13-5/18/05 was identified as an equipment blank. No contaminant concentrations were found in this blank with the following exceptions:

Equipment Blank ID	Analyte	Concentration
EB-13-5/18/05	pH Total dissolved solids Chloride Nitrate as N Sulfate	7.12 units 10.0 mg/L 0.15 mg/L 0.033 mg/L 0.35 mg/L

NASA JPL

Wet Chemistry - Data Qualification Summary - SDG 05-2772

No Sample Data Qualified in this SDG

NASA JPL

Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 05-2772

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 19, 2005
LDC Report Date: July 12, 2005
Matrix: Water
Parameters: Wet Chemistry
Validation Level: EPA Level III & IV
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 05-2788

Sample Identification

EB-14-5/19/05
MW-4-1**
MW-4-2
MW-4-3
MW-4-4
MW-4-5
MW-4-1MS
MW-4-1MSD
MW-4-5DUP

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 9 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 150.1 for pH, EPA Method 160.1 for Total Dissolved Solids, EPA Method 300.0 for Chloride, Nitrate as Nitrogen, and Sulfate, EPA Method 310.1 for Alkalinity, EPA Method 314.0 for Perchlorate, and EPA SW 846 Method 7196 for Hexavalent Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (October 2004) as there are no current guidelines for the methods stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification

All sample result verifications were acceptable for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

IX. Field Duplicates

No field duplicates were identified in this SDG.

X. Field Blanks

Sample EB-14-5/19/05 was identified as an equipment blank. No contaminant concentrations were found in this blank with the following exceptions:

Equipment Blank ID	Analyte	Concentration
EB-14-5/19/05	pH Chloride Nitrate as N Sulfate	6.30 units 0.15 mg/L 0.083 mg/L 0.37 mg/L

NASA JPL

Wet Chemistry - Data Qualification Summary - SDG 05-2788

No Sample Data Qualified in this SDG

NASA JPL

Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 05-2788

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 20, 2005
LDC Report Date: July 11, 2005
Matrix: Water
Parameters: Wet Chemistry
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 05-2812

Sample Identification

DUPE-8-2Q05
MW-5
MW-6
MW-16
MW-5MS
MW-5MSD
DUPE-8-2Q05DUP

Introduction

This data review covers 7 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 150.1 for pH, EPA Method 160.1 for Total Dissolved Solids, EPA Method 300.0 for Chloride, Nitrate as Nitrogen, and Sulfate, EPA Method 310.1 for Alkalinity, EPA Method 314.0 for Perchlorate, and EPA SW 846 Method 7196 for Hexavalent Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits with the following exceptions:

LCS ID (Associated Samples)	Analyte	LCS %R (Limits)	LCSD %R (Limits)	RPD (Limits)	Flag	A or P
LCS/LCSD (MW-5 MW-16)	Perchlorate	121 (80-120)	123 (80-120)	-	J (all detects)	P

VII. Sample Result Verification

Raw data were not reviewed for this SDG.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

IX. Field Duplicates

Samples DUPE-8-2Q05 and MW-6 were identified as field duplicates. No contaminant concentrations were detected in any of the samples with the following exceptions:

Analyte	Concentration (mg/L)		RPD
	DUPE-8-2Q05	MW-6	
Alkalinity	249	250	0
Total dissolved solids	744	801	7
Chloride	112	110	2
Nitrate as N	10.8	10.8	0
Sulfate	145	144	1

Analyte	Concentration (units)		RPD
	DUPE-8-2Q05	MW-6	
pH	6.71	6.70	0

Analyte	Concentration (ug/L)		RPD
	DUPE-8-2Q05	MW-6	
Perchlorate	2.1	2.9	32

X. Field Blanks

No field blanks were identified in this SDG.

NASA JPL**Wet Chemistry - Data Qualification Summary - SDG 05-2812**

SDG	Sample	Analyte	Flag	A or P	Reason
05-2812	MW-5 MW-16	Perchlorate	J (all detects)	P	Laboratory control samples (%R)

NASA JPL**Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 05-2812**

No Sample Data Qualified in this SDG

Laboratory Data Consultants, Inc.
Data Validation Report

Project/Site Name: NASA JPL
Collection Date: May 23, 2005
LDC Report Date: July 11, 2005
Matrix: Water
Parameters: Wet Chemistry
Validation Level: EPA Level III & IV
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2820

Sample Identification

MW-13**
MW-13MS
MW-13MSD

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 3 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 150.1 for pH, EPA Method 160.1 for Total Dissolved Solids, EPA Method 300.0 for Chloride, Nitrate as Nitrogen, and Sulfate, EPA Method 310.1 for Alkalinity, and EPA SW 846 Method 7196 for Hexavalent Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.

None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits with the following exceptions:

LCS ID (Associated Samples)	Analyte	LCS %R (Limits)	LCSD %R (Limits)	RPD (Limits)	Flag	A or P
LCS/LCSD (All samples in SDG 05-2820)	Perchlorate	121 (80-120)	123 (80-120)	-	J (all detects)	P

VII. Sample Result Verification

All sample result verifications were acceptable for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

IX. Field Duplicates

No field duplicates were identified in this SDG.

X. Field Blanks

No field blanks were identified in this SDG.

NASA JPL**Wet Chemistry - Data Qualification Summary - SDG 05-2820**

SDG	Sample	Analyte	Flag	A or P	Reason
05-2812	MW-13**	Perchlorate	J (all detects)	P	Laboratory control samples (%R)

NASA JPL**Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 05-2820**

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 24, 2005
LDC Report Date: July 12, 2005
Matrix: Water
Parameters: Wet Chemistry
Validation Level: EPA Level III & IV
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 05-2832

Sample Identification

DUPE-9-2Q05
MW-7
MW-8**
MW-10
DUPE-9-2Q05DUP
MW-7MS
MW-7MSD
MW-8MS
MW-8MSD

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 9 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 150.1 for pH, EPA Method 160.1 for Total Dissolved Solids, EPA Method 300.0 for Chloride, Nitrate as Nitrogen, and Sulfate, EPA Method 310.1 for Alkalinity, EPA Method 314.0 for Perchlorate, and EPA SW 846 Method 7196 for Hexavalent Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (October 2004) as there are no current guidelines for the methods stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification

All sample result verifications were acceptable for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

IX. Field Duplicates

Samples DUPE-9-2Q05 and MW-10 were identified as field duplicates. No contaminant concentrations were detected in any of the samples with the following exceptions:

Analyte	Concentration (mg/L)		RPD
	DUPE-9-2Q05	MW-10	
Alkalinity	253	250	1
Total dissolved solids	578	553	4
Hexavalent chromium	0.011	0.011	0
Chloride	43.6	42.8	2
Nitrate as N	13.0	12.7	2
Sulfate	98.9	96.7	2

Analyte	Concentration (units)		RPD
	DUPE-9-2Q05	MW-10	
pH	6.83	6.88	1

Analyte	Concentration (ug/L)		RPD
	DUPE-9-2Q05	MW-10	
Perchlorate	91.1	91.8	1

X. Field Blanks

No field blanks were identified in this SDG.

NASA JPL

Wet Chemistry - Data Qualification Summary - SDG 05-2832

No Sample Data Qualified in this SDG

NASA JPL

Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 05-2832

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 25, 2005
LDC Report Date: July 13, 2005
Matrix: Water
Parameters: Wet Chemistry
Validation Level: EPA Level III & IV
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 05-2848

Sample Identification

EB-19-5/25/05
MW-26-1
MW-26-2**
SB-1-2Q05
MW-26-1MS
MW-26-1MSD
MW-26-2DUP

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 7 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 150.1 for pH, EPA Method 160.1 for Total Dissolved Solids, EPA Method 300.0 for Chloride, Nitrate as Nitrogen, and Sulfate, EPA Method 310.1 for Alkalinity, EPA Method 314.0 for Perchlorate, and EPA SW 846 Method 7196 for Hexavalent Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (October 2004) as there are no current guidelines for the methods stated above.

A qualification summary table is provided at the end of this report if data has been qualified. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical or advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits with the following exceptions:

LCS ID (Associated Samples)	Analyte	LCS %R (Limits)	LCSD %R (Limits)	RPD (Limits)	Flag	A or P
LCS/LCSD (All samples in SDG 05-2848)	Perchlorate	121 (80-120)	123 (80-120)	-	J (all detects)	P

VII. Sample Result Verification

All sample result verifications were acceptable for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report if data has been qualified.

IX. Field Duplicates

No field duplicates were identified in this SDG.

X. Field Blanks

Sample EB-19-5/25/05 was identified as an equipment blank. No contaminant concentrations were found in this blank with the following exceptions:

Equipment Blank ID	Analyte	Concentration
EB-19-5/25/05	pH Chloride Nitrate as N	5.49 units 0.11 mg/L 0.033 mg/L

Sample SB-1-2Q05 was identified as a source blank. No contaminant concentrations were found in this blank with the following exceptions:

Source Blank ID	Analyte	Concentration
SB-1-2Q05	pH Chloride Nitrate as N Sulfate	6.14 units 0.11 mg/L 0.049 mg/L 0.39 mg/L

NASA JPL**Wet Chemistry - Data Qualification Summary - SDG 05-2848**

SDG	Sample	Analyte	Flag	A or P	Reason
05-2848	EB-19-5/25/05 MW-26-1 MW-26-2** SB-1-2Q05	Perchlorate	J (all detects)	P	Laboratory control samples (%R)

NASA JPL**Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 05-2848**

No Sample Data Qualified in this SDG